

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
 Data File : PP053298.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 17:20
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 20:23:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.648	106.0E6	88788352	55.138	59.025
2) SA Decachlor...	10.209	8.730	74904569	97117340	55.241	51.814
Target Compounds						
3) L1 AR-1016-1	5.577	4.748	32927773	31855921	529.302	567.982
4) L1 AR-1016-2	5.600	4.767	48531670	44384605	525.081	571.834
5) L1 AR-1016-3	5.662	4.946	29820188	23991430	526.891	571.367
6) L1 AR-1016-4	5.762	4.988	24144548	19066346	530.789	545.522
7) L1 AR-1016-5	6.058	5.205	23754698	25971886	524.793	571.814
31) L7 AR-1260-1	7.190	6.252	42456976	50661349	519.209	552.653
32) L7 AR-1260-2	7.448	6.442	48466286	62234829	523.348	540.831
33) L7 AR-1260-3	7.809	6.597	32773953	58944378	525.490	549.301
34) L7 AR-1260-4	8.036	7.074	39618416	45637698	528.857	551.620
35) L7 AR-1260-5	8.359	7.318	70484033	104.6E6	556.894	552.363



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
Data File : PP053298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2022 17:20
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 20:23:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 105993512
Conc: 55.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 88788352
Conc: 59.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: 0.006 min
Response: 74904569
Conc: 55.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.730 min
Delta R.T.: 0.002 min
Response: 97117340
Conc: 51.81 ng/ml



#3 AR-1016-1


R.T.: 5.577 min
Delta R.T.: 0.000 min
Response: 32927773
Conc: 529.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 31855921
Conc: 567.98 ng/ml




#4 AR-1016-2

R.T.: 5.600 min
Delta R.T.: 0.001 min
Response: 48531670
Conc: 525.08 ng/ml



#4 AR-1016-2


R.T.: 4.767 min
Delta R.T.: 0.000 min
Response: 44384605
Conc: 571.83 ng/ml



#5 AR-1016-3


R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 29820188
Conc: 526.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 23991430
Conc: 571.37 ng/ml




#6 AR-1016-4

R.T.: 5.762 min
Delta R.T.: 0.001 min
Response: 24144548
Conc: 530.79 ng/ml



#6 AR-1016-4


R.T.: 4.988 min
Delta R.T.: 0.000 min
Response: 19066346
Conc: 545.52 ng/ml



#7 AR-1016-5


R.T.: 6.058 min
Delta R.T.: 0.002 min
Response: 23754698
Conc: 524.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.205 min
Delta R.T.: 0.000 min
Response: 25971886
Conc: 571.81 ng/ml




#31 AR-1260-1

R.T.: 7.190 min
Delta R.T.: 0.002 min
Response: 42456976
Conc: 519.21 ng/ml



#31 AR-1260-1


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 50661349
Conc: 552.65 ng/ml



#32 AR-1260-2


R.T.: 7.448 min
Delta R.T.: 0.002 min
Response: 48466286
Conc: 523.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 62234829
Conc: 540.83 ng/ml




#33 AR-1260-3

R.T.: 7.809 min
Delta R.T.: 0.001 min
Response: 32773953
Conc: 525.49 ng/ml



#33 AR-1260-3


R.T.: 6.597 min
Delta R.T.: 0.000 min
Response: 58944378
Conc: 549.30 ng/ml



#34 AR-1260-4


R.T.: 8.036 min
Delta R.T.: 0.002 min
Response: 39618416
Conc: 528.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.074 min
Delta R.T.: 0.000 min
Response: 45637698
Conc: 551.62 ng/ml



#35 AR-1260-5

R.T.: 8.359 min
Delta R.T.: 0.002 min
Response: 70484033
Conc: 556.89 ng/ml



#35 AR-1260-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 104592393
Conc: 552.36 ng/ml