

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111124\
 Data File : PP068396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2024 09:57
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 12:31:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 11:04:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.760	4.051	50070884	49710169	51.587	51.338
2) SA Decachlor...	10.678	9.210	64468878	56017442	51.514	49.521
Target Compounds						
16) L4 AR-1242-1	5.925	5.158	14943792	14001218	500.236	503.826
17) L4 AR-1242-2	5.947	5.178	22052302	19367980	500.029	501.435
18) L4 AR-1242-3	6.011	5.359	14022591	10728880	503.352	504.175
19) L4 AR-1242-4	6.109	5.442	11229992	11346509	513.254	497.792
20) L4 AR-1242-5	6.843	5.971	11381441	13416358	468.750	488.732



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111124\
Data File : PP068396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2024 09:57
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 12:31:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 11:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.760 min
Delta R.T.: 0.003 min
Response: 50070884
Conc: 51.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 49710169
Conc: 51.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.678 min
Delta R.T.: 0.006 min
Response: 64468878
Conc: 51.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.210 min
Delta R.T.: 0.000 min
Response: 56017442
Conc: 49.52 ng/ml



#16 AR-1242-1


R.T.: 5.925 min
Delta R.T.: 0.000 min
Response: 14943792
Conc: 500.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.158 min
Delta R.T.: 0.000 min
Response: 14001218
Conc: 503.83 ng/ml




#17 AR-1242-2

R.T.: 5.947 min
Delta R.T.: 0.000 min
Response: 22052302
Conc: 500.03 ng/ml



#17 AR-1242-2


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 19367980
Conc: 501.44 ng/ml



#18 AR-1242-3


R.T.: 6.011 min
Delta R.T.: 0.000 min
Response: 14022591
Conc: 503.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.359 min
Delta R.T.: 0.000 min
Response: 10728880
Conc: 504.18 ng/ml




#19 AR-1242-4

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 11229992
Conc: 513.25 ng/ml



#19 AR-1242-4


R.T.: 5.442 min
Delta R.T.: 0.000 min
Response: 11346509
Conc: 497.79 ng/ml



#20 AR-1242-5

R.T.: 6.843 min
Delta R.T.: 0.001 min
Response: 11381441
Conc: 468.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.971 min
Delta R.T.: 0.000 min
Response: 13416358
Conc: 488.73 ng/ml