

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111124\
 Data File : PP068427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2024 19:28
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 02:07:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 11:04:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.759	4.051	49793716	49099687	51.301	50.708
2) SA Decachlor...	10.678	9.212	59250100	54861427	47.344	48.499
Target Compounds						
21) L5 AR-1248-1	5.924	5.158	11709462	11048901	518.934	505.730
22) L5 AR-1248-2	6.197	5.398	17001744	16011626	505.844	500.848
23) L5 AR-1248-3	6.403	5.442	18204683	16690247	498.544	498.094
24) L5 AR-1248-4	6.803	5.617	19429747	19483233	454.118	501.773
25) L5 AR-1248-5	6.843	6.014	18534965	18212429	446.204	491.321



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111124\
Data File : PP068427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2024 19:28
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 02:07:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 11:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.759 min
Delta R.T.: 0.002 min
Response: 49793716
Conc: 51.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 49099687
Conc: 50.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.678 min
Delta R.T.: 0.006 min
Response: 59250100
Conc: 47.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.212 min
Delta R.T.: 0.002 min
Response: 54861427
Conc: 48.50 ng/ml



#21 AR-1248-1


R.T.: 5.924 min
Delta R.T.: 0.003 min
Response: 11709462
Conc: 518.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.158 min
Delta R.T.: 0.000 min
Response: 11048901
Conc: 505.73 ng/ml




#22 AR-1248-2

R.T.: 6.197 min
Delta R.T.: 0.002 min
Response: 17001744
Conc: 505.84 ng/ml



#22 AR-1248-2


R.T.: 5.398 min
Delta R.T.: 0.000 min
Response: 16011626
Conc: 500.85 ng/ml



#23 AR-1248-3


R.T.: 6.403 min
Delta R.T.: 0.003 min
Response: 18204683
Conc: 498.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.442 min
Delta R.T.: 0.000 min
Response: 16690247
Conc: 498.09 ng/ml




#24 AR-1248-4

R.T.: 6.803 min
Delta R.T.: 0.004 min
Response: 19429747
Conc: 454.12 ng/ml



#24 AR-1248-4


R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 19483233
Conc: 501.77 ng/ml



#25 AR-1248-5

R.T.: 6.843 min
Delta R.T.: 0.003 min
Response: 18534965
Conc: 446.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.014 min
Delta R.T.: 0.000 min
Response: 18212429
Conc: 491.32 ng/ml