

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111218\
 Data File : PP028032.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Nov 2018 22:33
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 11/13/2018 4:34:21 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 06:10:39 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 17:06:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.088	12.202	197.2E6	317.6E6	757.504	748.386
Target Compounds							
1) T	Dalapon	3.338	3.008	425.1E6	375.2E6	677.246	641.762
2) T	3,5-DICHL...	10.501	10.292	359.1E6	360.5E6	669.745	696.833
3) T	4-Nitroph...	11.664	11.345	351.2E6	157.0E6	667.183	690.430
5) T	DICAMBA	12.410	12.512	1155.9E6	1177.6E6	698.587	701.218
6) T	MCP	12.753	12.704	33271478	61892034	71.714	74.554
7) T	MCPA	12.996	13.093	52164825	95406413	71.765	73.805
8) T	DICHLORPROP	13.606	13.626	265.2E6	286.8E6	619.797m	696.330
9) T	2,4-D	13.958	14.113	276.8E6	340.1E6	720.804	703.389
10) T	Pentachlo...	14.376	14.726	5856.7E6	4524.7E6	699.292	701.866
11) T	2,4,5-TP ...	15.239	15.295	2116.7E6	1714.7E6	703.031	698.484
12) T	2,4,5-T	15.628	15.936	1942.8E6	1513.4E6	700.523	696.361
13) T	2,4-DB	16.392	16.538	164.1E6	185.1E6	723.078	701.142
14) T	DINOSEB	17.880	16.930	1519.6E6	1202.1E6	669.557	668.670
15) T	Picloram	17.643	18.308	3335.9E6	2468.2E6	705.942	693.797
16) T	DCPA	18.245	18.191	3154.4E6	2218.7E6	697.670	690.247

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111218\
Data File : PP028032.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Nov 2018 22:33
Operator : EI\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
Client Sampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
11/13/2018 4:34:21 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 06:10:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 17:06:51 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

