

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111320\
 Data File : PP031259.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Nov 2020 11:45
 Operator : DD\AJ
 Sample : L4706-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-30706-30707-40854

Manual Integrations
 APPROVED

Ankita
 11/17/2020 8:58:16 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 13:55:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 05 13:03:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.818f	4.091f	1172432	1226943	17.554m	18.459
2) SA Decachlor...	10.723f	9.428f	706827	641716	17.678m	15.368

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111320\
Data File : PP031259.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Nov 2020 11:45
Operator : DD\AJ
Sample : L4706-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
COMP-30706-30707-40854


Manual Integrations
APPROVED

Ankita
11/17/2020 8:58:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 13:55:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 05 13:03:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.818 min
Delta R.T.: 0.017 min
Response: 1172432
Conc: 17.55 ng/ml m

Instrument :
ECD_P
ClientSampleId :
COMP-30706-30707-40854


Manual Integrations
APPROVED

Ankita
11/17/2020 8:58:16 AM




#1 Tetrachloro-m-xylene

R.T.: 4.091 min
Delta R.T.: 0.016 min
Response: 1226943
Conc: 18.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.723 min
Delta R.T.: 0.036 min
Response: 706827
Conc: 17.68 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.428 min
Delta R.T.: 0.025 min
Response: 641716
Conc: 15.37 ng/ml