

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111323\
 Data File : PP061632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 11:41
 Operator : YP\AJ
 Sample : 05353-03 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 EDA836K-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2023
 Supervised By :Ankita Jodhani 11/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 12:26:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 13 03:41:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.772	2101490	3614272	1.707m	1.749
2) SA Decachlor...	10.341	8.924	2308413	3608008	2.693	1.984 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111323\
Data File : PP061632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 11:41
Operator : YP\AJ
Sample : 05353-03 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
EDA836K-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 12:26:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 13 03:41:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.504 min
Delta R.T.: -0.005 min
Response: 2101490
Conc: 1.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
EDA836K-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023




#1 Tetrachloro-m-xylene

R.T.: 3.772 min
Delta R.T.: -0.004 min
Response: 3614272
Conc: 1.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.341 min
Delta R.T.: -0.008 min
Response: 2308413
Conc: 2.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.924 min
Delta R.T.: -0.010 min
Response: 3608008
Conc: 1.98 ng/ml