

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111323\
 Data File : PP061636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 12:47
 Operator : YP\AJ
 Sample : 05353-07 10X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC260548-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2023
 Supervised By :Ankita Jodhani 11/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 13:18:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 13 03:41:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.775	2038255	3316568	1.656	1.605
2) SA Decachlor...	10.345	8.927	2013754	3379322	2.349	1.858
Target Compounds						
31) L7 AR-1260-1	7.300	6.409	592692	1487661	11.768	12.418
32) L7 AR-1260-2	7.558	6.599	864268	2462417	16.091m	17.777
33) L7 AR-1260-3	7.919	6.756	487104	2172874	12.790	16.260 #
34) L7 AR-1260-4	8.145	7.234	675694	1360303	14.919m	13.863
35) L7 AR-1260-5	8.473	7.476	1428691	3103380	17.844	14.337

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111323\
Data File : PP061636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 12:47
Operator : YP\AJ
Sample : 05353-07 10X
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BC260548-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 13:18:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 13 03:41:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.507 min
Delta R.T.: -0.002 min
Response: 2038255
Conc: 1.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC260548-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023




#1 Tetrachloro-m-xylene

R.T.: 3.775 min
Delta R.T.: -0.002 min
Response: 3316568
Conc: 1.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.345 min
Delta R.T.: -0.004 min
Response: 2013754
Conc: 2.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.927 min
Delta R.T.: -0.008 min
Response: 3379322
Conc: 1.86 ng/ml




#31 AR-1260-1

R.T.: 7.300 min
Delta R.T.: -0.002 min
Response: 592692
Conc: 11.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC260548-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023




#31 AR-1260-1

R.T.: 6.409 min
Delta R.T.: -0.005 min
Response: 1487661
Conc: 12.42 ng/ml




#32 AR-1260-2

R.T.: 7.558 min
Delta R.T.: -0.002 min
Response: 864268
Conc: 16.09 ng/ml m



#32 AR-1260-2

R.T.: 6.599 min
Delta R.T.: -0.004 min
Response: 2462417
Conc: 17.78 ng/ml




#33 AR-1260-3

R.T.: 7.919 min
Delta R.T.: -0.001 min
Response: 487104
Conc: 12.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC260548-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023




#33 AR-1260-3

R.T.: 6.756 min
Delta R.T.: -0.004 min
Response: 2172874
Conc: 16.26 ng/ml




#34 AR-1260-4

R.T.: 8.145 min
Delta R.T.: -0.003 min
Response: 675694
Conc: 14.92 ng/ml m



#34 AR-1260-4

R.T.: 7.234 min
Delta R.T.: -0.005 min
Response: 1360303
Conc: 13.86 ng/ml




#35 AR-1260-5

R.T.: 8.473 min
Delta R.T.: -0.001 min
Response: 1428691
Conc: 17.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC260548-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2023
Supervised By :Ankita Jodhani 11/14/2023



#35 AR-1260-5

R.T.: 7.476 min
Delta R.T.: -0.005 min
Response: 3103380
Conc: 14.34 ng/ml