

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111521\
 Data File : PP041045.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Nov 2021 1:25
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 04:47:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	21.089	21.249 E3	-0.8	98	0.00
2 SA Decachlorobiphenyl	21.844	21.336 E3	2.3	99	0.00
26 L6 AR-1254-1	772.268	773.562	-0.2	99	0.00
27 L6 AR-1254-2	1.236	1.218 E3	1.5	98	0.00
28 L6 AR-1254-3	1.339	1.336 E3	0.2	99	0.00
29 L6 AR-1254-4	954.334	968.412	-1.5	99	0.00
30 L6 AR-1254-5	1.063	1.088 E3	-2.4	99	0.00

Signal #2

1 SA Tetrachloro-m-xylene	27.791	29.352 E3	-5.6	106	0.00
2 SA Decachlorobiphenyl	14.335	13.895 E3	3.1	98	0.00
26 L6 AR-1254-1	1.139	1.114 E3	2.2	100	0.00
27 L6 AR-1254-2	980.538	949.583	3.2	100	0.00
28 L6 AR-1254-3	1.446	1.407 E3	2.7	100	0.00
29 L6 AR-1254-4	829.498	823.918	0.7	100	0.00
30 L6 AR-1254-5	1.166	1.154 E3	1.0	99	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	703.065	0.000	100.0#	0#	-6.07#
4 L1 AR-1016-2	1.111	0.000 E3	100.0#	0#	-6.10#
5 L1 AR-1016-3	711.462	0.000	100.0#	0#	-6.16#
6 L1 AR-1016-4	564.071	0.000	100.0#	0#	-6.27#
7 L1 AR-1016-5	553.019	0.000	100.0#	0#	-6.58#
8 L2 AR-1221-1	226.545	0.000	100.0#	0#	-5.01#
9 L2 AR-1221-2	154.802	0.000	100.0#	0#	-5.11#
10 L2 AR-1221-3	534.948	0.000	100.0#	0#	-5.19#
11 L3 AR-1232-1	443.019	0.000	100.0#	0#	-5.19#
12 L3 AR-1232-2	230.774	0.000	100.0#	0#	-5.78#
13 L3 AR-1232-3	427.476	0.000	100.0#	0#	-6.10#
14 L3 AR-1232-4	203.216	0.000	100.0#	0#	-6.27#
15 L3 AR-1232-5	141.167	0.000	100.0#	0#	-6.37#
16 L4 AR-1242-1	516.921	0.000	100.0#	0#	-6.07#
17 L4 AR-1242-2	816.855	0.000	100.0#	0#	-6.10#
18 L4 AR-1242-3	525.532	0.000	100.0#	0#	-6.16#
19 L4 AR-1242-4	403.559	0.000	100.0#	0#	-6.27#
20 L4 AR-1242-5	447.019	0.000	100.0#	0#	-7.05#
21 L5 AR-1248-1	371.544	0.000	100.0#	0#	-6.07#
22 L5 AR-1248-2	554.503	0.000	100.0#	0#	-6.37#
23 L5 AR-1248-3	673.378	0.000	100.0#	0#	-6.58#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111521\
 Data File : PP041045.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Nov 2021 1:25
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 04:47:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	746.078	0.000	100.0#	0#	-7.01#
25 L5 AR-1248-5	756.076	0.000	100.0#	0#	-7.05#
31 L7 AR-1260-1	0.986	0.000 E3	100.0#	0#	-7.76#
32 L7 AR-1260-2	1.182	0.000 E3	100.0#	0#	-8.02#
33 L7 AR-1260-3	907.809	0.000	100.0#	0#	-8.39#
34 L7 AR-1260-4	1.081	0.000 E3	100.0#	0#	-8.62#
35 L7 AR-1260-5	2.084	0.000 E3	100.0#	0#	-8.93#
36 L8 AR-1262-1	1.195	0.000 E3	100.0#	0#	-8.38#
37 L8 AR-1262-2	2.109	0.000 E3	100.0#	0#	-8.93#
38 L8 AR-1262-3	1.009	0.000 E3	100.0#	0#	-9.23#
39 L8 AR-1262-4	667.432	0.000	100.0#	0#	-9.31#
40 L8 AR-1262-5	823.678	0.000	100.0#	0#	-9.95#
41 L9 AR-1268-1	2.721	0.000 E3	100.0#	0#	-9.22#
42 L9 AR-1268-2	2.573	0.000 E3	100.0#	0#	-9.32#
43 L9 AR-1268-3	2.190	0.000 E3	100.0#	0#	-9.53#
44 L9 AR-1268-4	959.568	0.000	100.0#	0#	-9.95#
45 L9 AR-1268-5	7.066	0.000 E3	100.0#	0#	-10.34#

Signal #2

3 L1 AR-1016-1	848.658	0.000	100.0#	0#	-5.01#
4 L1 AR-1016-2	1.229	0.000 E3	100.0#	0#	-5.03#
5 L1 AR-1016-3	673.815	0.000	100.0#	0#	-5.22#
6 L1 AR-1016-4	543.830	0.000	100.0#	0#	-5.27#
7 L1 AR-1016-5	666.238	0.000	100.0#	0#	-5.50#
8 L2 AR-1221-1	319.247	0.000	100.0#	0#	-4.02#
9 L2 AR-1221-2	253.563	0.000	100.0#	0#	-4.12#
10 L2 AR-1221-3	751.344	0.000	100.0#	0#	-4.20#
11 L3 AR-1232-1	598.907	0.000	100.0#	0#	-4.20#
12 L3 AR-1232-2	497.071	0.000	100.0#	0#	-5.03#
13 L3 AR-1232-3	267.749	0.000	100.0#	0#	-5.22#
14 L3 AR-1232-4	234.597	0.000	100.0#	0#	-5.32#
15 L3 AR-1232-5	245.910	0.000	100.0#	0#	-5.50#
16 L4 AR-1242-1	634.738	0.000	100.0#	0#	-5.01#
17 L4 AR-1242-2	919.616	0.000	100.0#	0#	-5.03#
18 L4 AR-1242-3	507.642	0.000	100.0#	0#	-5.22#
19 L4 AR-1242-4	490.885	0.000	100.0#	0#	-5.31#
20 L4 AR-1242-5	539.909	0.000	100.0#	0#	-5.88#
21 L5 AR-1248-1	483.076	0.000	100.0#	0#	-5.01#
22 L5 AR-1248-2	678.443	0.000	100.0#	0#	-5.27#
23 L5 AR-1248-3	697.777	0.000	100.0#	0#	-5.31#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111521\
 Data File : PP041045.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Nov 2021 1:25
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 04:47:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	800.643	0.000	100.0#	0#	-5.50#
25 L5 AR-1248-5	723.962	0.000	100.0#	0#	-5.92#
31 L7 AR-1260-1	957.337	0.000	100.0#	0#	-6.59#
32 L7 AR-1260-2	1.085	0.000 E3	100.0#	0#	-6.79#
33 L7 AR-1260-3	1003.377	0.000	100.0#	0#	-6.95#
34 L7 AR-1260-4	820.184	0.000	100.0#	0#	-7.43#
35 L7 AR-1260-5	1.745	0.000 E3	100.0#	0#	-7.68#
36 L8 AR-1262-1	592.350	0.000	100.0#	0#	-7.12#
37 L8 AR-1262-2	1.655	0.000 E3	100.0#	0#	-7.68#
38 L8 AR-1262-3	720.317	0.000	100.0#	0#	-7.96#
39 L8 AR-1262-4	1.286	0.000 E3	100.0#	0#	-8.03#
40 L8 AR-1262-5	600.429	0.000	100.0#	0#	-8.53#
41 L9 AR-1268-1	2.083	0.000 E3	100.0#	0#	-7.96#
42 L9 AR-1268-2	1.905	0.000 E3	100.0#	0#	-8.03#
43 L9 AR-1268-3	1.624	0.000 E3	100.0#	0#	-8.23#
44 L9 AR-1268-4	702.938	0.000	100.0#	0#	-8.53#
45 L9 AR-1268-5	4.646	0.000 E3	100.0#	0#	-8.80#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0