

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111523\
 Data File : PP061697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2023 18:52
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP111523AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 00:26:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 00:24:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.727	68815684	110.9E6	51.433	50.516
2) SA Decachlor...	10.240	8.859	52086528	83734384	50.954	51.051
Target Compounds						
21) L5 AR-1248-1	5.626	4.842	14400590	20677202	509.860	504.631
22) L5 AR-1248-2	5.901	5.084	23206625	30748461	514.795	507.521
23) L5 AR-1248-3	6.106	5.126	26100326	31882787	531.116	505.807
24) L5 AR-1248-4	6.511	5.302	26299085	36029783	508.921	504.291
25) L5 AR-1248-5	6.550	5.660	27437917	45676961	517.884	505.184

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111523\
Data File : PP061697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2023 18:52
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP111523AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 00:26:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 00:24:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

