

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111820\
 Data File : PP031364.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Nov 2020 16:10
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 01:52:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 18 01:50:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.795	4.069	2332078	2354127	57.244	56.308
2)	SA Decachlor...	10.669	9.380	1570294	1818739	51.126	54.315
Target Compounds							
3)	L1 AR-1016-1	6.098	5.326	703991	710359	559.521	557.717
4)	L1 AR-1016-2	6.122	5.346	1043494	1036298	554.883	552.895
5)	L1 AR-1016-3	6.187	5.538	643721	565453	562.380	565.919
6)	L1 AR-1016-4	6.293	5.589	531021	423027	552.104	574.439
7)	L1 AR-1016-5	6.607	5.817	492068	497216	561.417	511.519
8)	L2 AR-1221-1	5.035	4.325	71288	69338	151.853	151.698
9)	L2 AR-1221-2	5.138	4.425	108370	105130	311.660	297.709
10)	L2 AR-1221-3	5.223	4.514	405934	401131	380.771	370.530
11)	L3 AR-1232-1	5.223	4.514	405934	401131	469.288	466.497
12)	L3 AR-1232-2	5.806	5.346	547608	1036298	1252.370	1259.576
13)	L3 AR-1232-3	6.122	5.538	1043494	565453	1255.018	1304.779
14)	L3 AR-1232-4	6.293	5.634	531021	516455	1263.112	1458.937
15)	L3 AR-1232-5	6.393	5.817	380902	497216	1504.259	1249.323
16)	L4 AR-1242-1	6.098	5.326	703991	710359	734.535	742.458
17)	L4 AR-1242-2	6.122	5.346	1043494	1036298	734.382	736.548
18)	L4 AR-1242-3	6.187	5.538	643721	565453	725.012	731.161
19)	L4 AR-1242-4	6.293	5.634	531021	516455	713.396	731.953
20)	L4 AR-1242-5	7.072	6.197	74893	392133	101.546	443.694 #
21)	L5 AR-1248-1	6.098	5.326	703991	710359	946.279	945.803
22)	L5 AR-1248-2	6.393	5.589	380902	423027	405.190	420.561
23)	L5 AR-1248-3	6.607	5.634	492068	516455	419.999	482.730
24)	L5 AR-1248-4	7.033	5.817	89755	497216	71.079	388.734 #
25)	L5 AR-1248-5	7.072	6.241	74893	66186	59.620	52.694
26)	L6 AR-1254-1	7.004	6.197	307040	392133	239.035	206.058
27)	L6 AR-1254-2	7.232	6.357	331537	352190	171.733	215.160 #
28)	L6 AR-1254-3	7.611	6.796	216428	693031	99.839	253.844 #
29)	L6 AR-1254-4	7.904	7.017	148885	87759	98.607	56.473 #
30)	L6 AR-1254-5	8.331	7.444	1067664	1198631	653.100	495.756
31)	L7 AR-1260-1	7.779	6.914	747418	890078	535.164	532.692
32)	L7 AR-1260-2	8.043	7.110	878992	1056584	528.225	519.193
33)	L7 AR-1260-3	8.409	7.266	698285	997426	546.607	516.958
34)	L7 AR-1260-4	8.637	7.748	811928	868013	531.109	538.242
35)	L7 AR-1260-5	8.951	7.995	1616885	2097523	513.040	530.916
36)	L8 AR-1262-1	8.409	7.486	698285	932707	386.713	409.264
37)	L8 AR-1262-2	8.951	7.995	1616885	2097523	479.112	509.989
38)	L8 AR-1262-3	9.265	8.282	1044974	493431	447.085	309.582 #
39)	L8 AR-1262-4	9.338	8.345	327157	1550013	173.864	495.087 #
40)	L8 AR-1262-5	9.965	8.843	510752	617510	400.893	437.578
41)	L9 AR-1268-1	9.265	8.282	1044974	493431	247.607	100.251 #
42)	L9 AR-1268-2	9.338	8.345	327157	1550013	79.212	318.782 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111820\
 Data File : PP031364.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Nov 2020 16:10
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 01:52:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 18 01:50:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	0.000	8.555	0	27248	N.D.	6.542 #
44) L9	AR-1268-4	9.965	8.843	510752	617510	353.438	370.521
45) L9	AR-1268-5	10.354	9.130	146665	168509	13.016	12.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111820\
Data File : PP031364.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Nov 2020 16:10
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 01:52:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 18 01:50:50 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.795 min
Delta R.T.: -0.003 min
Response: 2332078
Conc: 57.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.069 min
Delta R.T.: -0.003 min
Response: 2354127
Conc: 56.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.669 min
Delta R.T.: -0.011 min
Response: 1570294
Conc: 51.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.380 min
Delta R.T.: -0.008 min
Response: 1818739
Conc: 54.32 ng/ml



#3 AR-1016-1


R.T.: 6.098 min
Delta R.T.: -0.004 min
Response: 703991
Conc: 559.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.326 min
Delta R.T.: -0.004 min
Response: 710359
Conc: 557.72 ng/ml




#4 AR-1016-2

R.T.: 6.122 min
Delta R.T.: -0.004 min
Response: 1043494
Conc: 554.88 ng/ml



#4 AR-1016-2


R.T.: 5.346 min
Delta R.T.: -0.005 min
Response: 1036298
Conc: 552.89 ng/ml



#5 AR-1016-3


R.T.: 6.187 min
Delta R.T.: -0.004 min
Response: 643721
Conc: 562.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.538 min
Delta R.T.: -0.005 min
Response: 565453
Conc: 565.92 ng/ml




#6 AR-1016-4

R.T.: 6.293 min
Delta R.T.: -0.004 min
Response: 531021
Conc: 552.10 ng/ml



#6 AR-1016-4


R.T.: 5.589 min
Delta R.T.: -0.005 min
Response: 423027
Conc: 574.44 ng/ml



#7 AR-1016-5


R.T.: 6.607 min
Delta R.T.: -0.004 min
Response: 492068
Conc: 561.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.817 min
Delta R.T.: -0.006 min
Response: 497216
Conc: 511.52 ng/ml




#8 AR-1221-1

R.T.: 5.035 min
Delta R.T.: -0.008 min
Response: 71288
Conc: 151.85 ng/ml



#8 AR-1221-1


R.T.: 4.325 min
Delta R.T.: -0.005 min
Response: 69338
Conc: 151.70 ng/ml



#9 AR-1221-2


R.T.: 5.138 min
Delta R.T.: -0.002 min
Response: 108370
Conc: 311.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.425 min
Delta R.T.: -0.004 min
Response: 105130
Conc: 297.71 ng/ml




#10 AR-1221-3

R.T.: 5.223 min
Delta R.T.: -0.003 min
Response: 405934
Conc: 380.77 ng/ml



#10 AR-1221-3


R.T.: 4.514 min
Delta R.T.: -0.005 min
Response: 401131
Conc: 370.53 ng/ml



#11 AR-1232-1


R.T.: 5.223 min
Delta R.T.: -0.003 min
Response: 405934
Conc: 469.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.514 min
Delta R.T.: -0.004 min
Response: 401131
Conc: 466.50 ng/ml




#12 AR-1232-2

R.T.: 5.806 min
Delta R.T.: -0.003 min
Response: 547608
Conc: 1252.37 ng/ml



#12 AR-1232-2


R.T.: 5.346 min
Delta R.T.: -0.004 min
Response: 1036298
Conc: 1259.58 ng/ml



#13 AR-1232-3


R.T.: 6.122 min
Delta R.T.: -0.003 min
Response: 1043494
Conc: 1255.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.538 min
Delta R.T.: -0.004 min
Response: 565453
Conc: 1304.78 ng/ml




#14 AR-1232-4

R.T.: 6.293 min
Delta R.T.: -0.003 min
Response: 531021
Conc: 1263.11 ng/ml



#14 AR-1232-4


R.T.: 5.634 min
Delta R.T.: -0.004 min
Response: 516455
Conc: 1458.94 ng/ml



#15 AR-1232-5


R.T.: 6.393 min
Delta R.T.: -0.004 min
Response: 380902
Conc: 1504.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.817 min
Delta R.T.: -0.006 min
Response: 497216
Conc: 1249.32 ng/ml




#16 AR-1242-1

R.T.: 6.098 min
Delta R.T.: -0.004 min
Response: 703991
Conc: 734.54 ng/ml



#16 AR-1242-1


R.T.: 5.326 min
Delta R.T.: -0.004 min
Response: 710359
Conc: 742.46 ng/ml



#17 AR-1242-2


R.T.: 6.122 min
Delta R.T.: -0.004 min
Response: 1043494
Conc: 734.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.346 min
Delta R.T.: -0.004 min
Response: 1036298
Conc: 736.55 ng/ml




#18 AR-1242-3

R.T.: 6.187 min
Delta R.T.: -0.004 min
Response: 643721
Conc: 725.01 ng/ml



#18 AR-1242-3


R.T.: 5.538 min
Delta R.T.: -0.004 min
Response: 565453
Conc: 731.16 ng/ml



#19 AR-1242-4


R.T.: 6.293 min
Delta R.T.: -0.004 min
Response: 531021
Conc: 713.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.634 min
Delta R.T.: -0.004 min
Response: 516455
Conc: 731.95 ng/ml




#20 AR-1242-5

R.T.: 7.072 min
Delta R.T.: -0.003 min
Response: 74893
Conc: 101.55 ng/ml



#20 AR-1242-5


R.T.: 6.197 min
Delta R.T.: -0.005 min
Response: 392133
Conc: 443.69 ng/ml



#21 AR-1248-1


R.T.: 6.098 min
Delta R.T.: -0.003 min
Response: 703991
Conc: 946.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.326 min
Delta R.T.: -0.004 min
Response: 710359
Conc: 945.80 ng/ml




#22 AR-1248-2

R.T.: 6.393 min
Delta R.T.: -0.003 min
Response: 380902
Conc: 405.19 ng/ml



#22 AR-1248-2


R.T.: 5.589 min
Delta R.T.: -0.004 min
Response: 423027
Conc: 420.56 ng/ml



#23 AR-1248-3


R.T.: 6.607 min
Delta R.T.: -0.004 min
Response: 492068
Conc: 420.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.634 min
Delta R.T.: -0.004 min
Response: 516455
Conc: 482.73 ng/ml




#24 AR-1248-4

R.T.: 7.033 min
Delta R.T.: -0.004 min
Response: 89755
Conc: 71.08 ng/ml



#24 AR-1248-4


R.T.: 5.817 min
Delta R.T.: -0.005 min
Response: 497216
Conc: 388.73 ng/ml



#25 AR-1248-5


R.T.: 7.072 min
Delta R.T.: -0.003 min
Response: 74893
Conc: 59.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 6.241 min
Delta R.T.: -0.004 min
Response: 66186
Conc: 52.69 ng/ml




#26 AR-1254-1

R.T.: 7.004 min
Delta R.T.: -0.004 min
Response: 307040
Conc: 239.04 ng/ml



#26 AR-1254-1


R.T.: 6.197 min
Delta R.T.: -0.004 min
Response: 392133
Conc: 206.06 ng/ml



#27 AR-1254-2


R.T.: 7.232 min
Delta R.T.: -0.005 min
Response: 331537
Conc: 171.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 352190
Conc: 215.16 ng/ml




#28 AR-1254-3

R.T.: 7.611 min
Delta R.T.: -0.004 min
Response: 216428
Conc: 99.84 ng/ml



#28 AR-1254-3


R.T.: 6.796 min
Delta R.T.: 0.014 min
Response: 693031
Conc: 253.84 ng/ml



#29 AR-1254-4


R.T.: 7.904 min
Delta R.T.: -0.005 min
Response: 148885
Conc: 98.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 7.017 min
Delta R.T.: -0.004 min
Response: 87759
Conc: 56.47 ng/ml




#30 AR-1254-5

R.T.: 8.331 min
Delta R.T.: -0.007 min
Response: 1067664
Conc: 653.10 ng/ml



#30 AR-1254-5


R.T.: 7.444 min
Delta R.T.: -0.006 min
Response: 1198631
Conc: 495.76 ng/ml



#31 AR-1260-1


R.T.: 7.779 min
Delta R.T.: -0.004 min
Response: 747418
Conc: 535.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.914 min
Delta R.T.: -0.006 min
Response: 890078
Conc: 532.69 ng/ml




#32 AR-1260-2

R.T.: 8.043 min
Delta R.T.: -0.006 min
Response: 878992
Conc: 528.23 ng/ml



#32 AR-1260-2


R.T.: 7.110 min
Delta R.T.: -0.006 min
Response: 1056584
Conc: 519.19 ng/ml



#33 AR-1260-3


R.T.: 8.409 min
Delta R.T.: -0.005 min
Response: 698285
Conc: 546.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 7.266 min
Delta R.T.: -0.005 min
Response: 997426
Conc: 516.96 ng/ml




#34 AR-1260-4

R.T.: 8.637 min
Delta R.T.: -0.006 min
Response: 811928
Conc: 531.11 ng/ml



#34 AR-1260-4


R.T.: 7.748 min
Delta R.T.: -0.006 min
Response: 868013
Conc: 538.24 ng/ml



#35 AR-1260-5


R.T.: 8.951 min
Delta R.T.: -0.006 min
Response: 1616885
Conc: 513.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.995 min
Delta R.T.: -0.006 min
Response: 2097523
Conc: 530.92 ng/ml




#36 AR-1262-1

R.T.: 8.409 min
Delta R.T.: -0.003 min
Response: 698285
Conc: 386.71 ng/ml



#36 AR-1262-1


R.T.: 7.486 min
Delta R.T.: -0.005 min
Response: 932707
Conc: 409.26 ng/ml



#37 AR-1262-2


R.T.: 8.951 min
Delta R.T.: -0.004 min
Response: 1616885
Conc: 479.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.995 min
Delta R.T.: -0.004 min
Response: 2097523
Conc: 509.99 ng/ml




#38 AR-1262-3

R.T.: 9.265 min
Delta R.T.: 0.009 min
Response: 1044974
Conc: 447.08 ng/ml



#38 AR-1262-3


R.T.: 8.282 min
Delta R.T.: -0.005 min
Response: 493431
Conc: 309.58 ng/ml



#39 AR-1262-4


R.T.: 9.338 min
Delta R.T.: -0.006 min
Response: 327157
Conc: 173.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.345 min
Delta R.T.: -0.006 min
Response: 1550013
Conc: 495.09 ng/ml




#40 AR-1262-5

R.T.: 9.965 min
Delta R.T.: -0.006 min
Response: 510752
Conc: 400.89 ng/ml



#40 AR-1262-5


R.T.: 8.843 min
Delta R.T.: -0.005 min
Response: 617510
Conc: 437.58 ng/ml



#41 AR-1268-1


R.T.: 9.265 min
Delta R.T.: 0.010 min
Response: 1044974
Conc: 247.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 8.282 min
Delta R.T.: -0.005 min
Response: 493431
Conc: 100.25 ng/ml




#42 AR-1268-2

R.T.: 9.338 min
Delta R.T.: -0.008 min
Response: 327157
Conc: 79.21 ng/ml



#42 AR-1268-2


R.T.: 8.345 min
Delta R.T.: -0.007 min
Response: 1550013
Conc: 318.78 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.562 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 8.555 min
Delta R.T.: -0.005 min
Response: 27248
Conc: 6.54 ng/ml




#44 AR-1268-4

R.T.: 9.965 min
Delta R.T.: -0.007 min
Response: 510752
Conc: 353.44 ng/ml



#44 AR-1268-4


R.T.: 8.843 min
Delta R.T.: -0.006 min
Response: 617510
Conc: 370.52 ng/ml



#45 AR-1268-5

R.T.: 10.354 min
Delta R.T.: -0.009 min
Response: 146665
Conc: 13.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 9.130 min
Delta R.T.: -0.006 min
Response: 168509
Conc: 12.91 ng/ml