

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP111820\  
 Data File : PP031390.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 18 Nov 2020 23:57  
 Operator : DD\AJ  
 Sample : L4778-11  
 Misc :  
 ALS Vial : 44 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 1814

Manual Integrations  
 APPROVED

Ankita  
 11/19/2020 1:39:46 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 19 09:27:47 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP111720.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Nov 18 01:50:50 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1 | Resp#2 | ng/ml | ng/ml   |
|-----------------------------|--------|-------|--------|--------|-------|---------|
| -----                       |        |       |        |        |       |         |
| System Monitoring Compounds |        |       |        |        |       |         |
| 1) SA Tetrachlo...          | 4.793  | 4.068 | 356193 | 148827 | 8.743 | 3.560m# |
| 2) SA Decachlor...          | 10.666 | 9.376 | 183363 | 333188 | 5.970 | 9.950 # |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP111820\  
Data File : PP031390.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 18 Nov 2020 23:57  
Operator : DD\AJ  
Sample : L4778-11  
Misc :  
ALS Vial : 44 Sample Multiplier: 1



Instrument :  
ECD\_P  
ClientSampleId :  
1814


Manual Integrations  
APPROVED

Ankita  
11/19/2020 1:39:46 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 19 09:27:47 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP111720.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Nov 18 01:50:50 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.793 min  
Delta R.T.: -0.003 min  
Response: 356193  
Conc: 8.74 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
1814


Manual Integrations  
APPROVED

Ankita  
11/19/2020 1:39:46 PM




#1 Tetrachloro-m-xylene

R.T.: 4.068 min  
Delta R.T.: -0.002 min  
Response: 148827  
Conc: 3.56 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.666 min  
Delta R.T.: -0.007 min  
Response: 183363  
Conc: 5.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.376 min  
Delta R.T.: -0.006 min  
Response: 333188  
Conc: 9.95 ng/ml