

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111822\
 Data File : PP053374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Nov 2022 15:05
 Operator : YP\AJ
 Sample : N5667-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/21/2022
 Supervised By :Ankita Jodhani 11/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 00:47:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.400	3.646	14103253	22245622	7.337m	14.789m#
2) SA Decachlor...	10.205	8.727	23642104	24876917	17.436	13.272

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111822\
Data File : PP053374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Nov 2022 15:05
Operator : YP\AJ
Sample : N5667-01
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
3


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/21/2022
Supervised By :Ankita Jodhani 11/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 00:47:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.400 min
Delta R.T.: -0.003 min
Response: 14103253
Conc: 7.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
3


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/21/2022
Supervised By :Ankita Jodhani 11/21/2022




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: -0.002 min
Response: 22245622
Conc: 14.79 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.205 min
Delta R.T.: 0.002 min
Response: 23642104
Conc: 17.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.727 min
Delta R.T.: -0.002 min
Response: 24876917
Conc: 13.27 ng/ml