

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111921\
 Data File : PP041093.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Nov 2021 9:14
 Operator : AJ\MA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 01:11:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
2) SA Decachlor...	10.624f	0.000	33643	0	1.540	N.D. #
Target Compounds						
20) L4 AR-1242-5	7.053	5.901f	14970	42473	33.487	78.667 #
24) L5 AR-1248-4	7.026	0.000	12636	0	16.936	N.D. #
25) L5 AR-1248-5	7.053	5.901f	14970	42473	19.799	58.668 #
26) L6 AR-1254-1	6.991	5.901f	123210	42473	159.543	37.299 #
28) L6 AR-1254-3	7.606f	6.456	38245	12009	28.562	8.302 #
29) L6 AR-1254-4	7.889	0.000	62706	0	65.706	N.D. #
32) L7 AR-1260-2	0.000	6.764f	0	36522	N.D.	33.660 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111921\
Data File : PP041093.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Nov 2021 9:14
Operator : AJ\MA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 01:11:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 4.757 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.760 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 10.624 min
Delta R.T.: -0.036 min
Response: 33643
Conc: 1.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T.: 9.026 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 7.053 min
Delta R.T.: 0.002 min
Response: 14970
Conc: 33.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE




#20 AR-1242-5

R.T.: 5.901 min
Delta R.T.: 0.023 min
Response: 42473
Conc: 78.67 ng/ml




#24 AR-1248-4

R.T.: 7.026 min
Delta R.T.: 0.015 min
Response: 12636
Conc: 16.94 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.498 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 7.053 min
Delta R.T.: 0.002 min
Response: 14970
Conc: 19.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 5.901 min
Delta R.T.: -0.019 min
Response: 42473
Conc: 58.67 ng/ml




#26 AR-1254-1

R.T.: 6.991 min
Delta R.T.: 0.012 min
Response: 123210
Conc: 159.54 ng/ml



#26 AR-1254-1


R.T.: 5.901 min
Delta R.T.: 0.024 min
Response: 42473
Conc: 37.30 ng/ml



#28 AR-1254-3


R.T.: 7.606 min
Delta R.T.: 0.017 min
Response: 38245
Conc: 28.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE




#28 AR-1254-3

R.T.: 6.456 min
Delta R.T.: 0.000 min
Response: 12009
Conc: 8.30 ng/ml




#29 AR-1254-4

R.T.: 7.889 min
Delta R.T.: 0.005 min
Response: 62706
Conc: 65.71 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.697 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 8.022 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEXANE



#32 AR-1260-2

R.T.: 6.764 min
Delta R.T.: -0.029 min
Response: 36522
Conc: 33.66 ng/ml