

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111921\
 Data File : PP041172.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Nov 2021 9:43
 Operator : AJ\MA
 Sample : PB140869BL
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB140869BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 01:37:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.755	3.759	431501	640174	20.461	23.036
2) SA Decachlor...	10.650	9.016	515530	334019	23.600	23.301

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111921\
Data File : PP041172.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Nov 2021 9:43
Operator : AJ\MA
Sample : PB140869BL
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB140869BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 01:37:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.755 min
Delta R.T.: -0.002 min
Response: 431501
Conc: 20.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB140869BL




#1 Tetrachloro-m-xylene

R.T.: 3.759 min
Delta R.T.: -0.002 min
Response: 640174
Conc: 23.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.650 min
Delta R.T.: -0.011 min
Response: 515530
Conc: 23.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: -0.010 min
Response: 334019
Conc: 23.30 ng/ml