

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112018\
 Data File : PP028122.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Nov 2018 22:29
 Operator : EI\MA
 Sample : J6006-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 RS-CON-WALL-1

Manual Integrations
 APPROVED

mohammad
 11/23/2018 10:52:01 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 07:47:46 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 17:06:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.094	12.205	191.3E6	183.6E6	734.892m	432.663 #
Target Compounds							
9) T	2,4-D	13.963	14.116	91070435	111.1E6	237.179	229.836

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112018\
Data File : PP028122.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Nov 2018 22:29
Operator : EI\MA
Sample : J6006-02
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleID :
RS-CON-WALL-1


Manual Integrations
APPROVED

mohammad
11/23/2018 10:52:01 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 07:47:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 17:06:51 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 12.094 min
Delta R.T.: 0.006 min
Response: 191297997
Conc: 734.89 ng/ml m

Instrument :
ECD_P
ClientSampleId :
RS-CON-WALL-1


Manual Integrations
APPROVED

mohammad
11/23/2018 10:52:01 AM




#4 2,4-DCAA

R.T.: 12.205 min
Delta R.T.: 0.004 min
Response: 183633817
Conc: 432.66 ng/ml



#9 2,4-D

R.T.: 13.963 min
Delta R.T.: 0.005 min
Response: 91070435
Conc: 237.18 ng/ml



#9 2,4-D

R.T.: 14.116 min
Delta R.T.: 0.000 min
Response: 111133658
Conc: 229.84 ng/ml