

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112020\
 Data File : PP031450.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Nov 2020 13:07
 Operator : DD\AJ
 Sample : L4807-01 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 72-11929-212

Manual Integrations
 APPROVED

mohammad
 11/23/2020 12:05:02 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 01:22:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 18 01:50:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.793	4.067	157861	158505	3.875	3.791
2)	SA Decachlor...	10.668	9.373	182322	207972	5.936	6.211
Target Compounds							
26)	L6 AR-1254-1	7.004	6.195	71612	79405	55.751m	41.726 #
27)	L6 AR-1254-2	7.232	6.353	97033	51474	50.262	31.446 #
28)	L6 AR-1254-3	7.611	6.773	131774	86919	60.788m	31.837 #
29)	L6 AR-1254-4	7.905	7.013	42569	33880	28.193m	21.802
30)	L6 AR-1254-5	8.330	7.440	100299	129759	61.354m	53.669

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112020\
Data File : PP031450.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Nov 2020 13:07
Operator : DD\AJ
Sample : L4807-01 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
72-11929-212


Manual Integrations
APPROVED

mohammad
11/23/2020 12:05:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 01:22:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 18 01:50:50 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.793 min
Delta R.T.: -0.003 min
Response: 157861
Conc: 3.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
72-11929-212


Manual Integrations
APPROVED

mohammad
11/23/2020 12:05:02 PM




#1 Tetrachloro-m-xylene

R.T.: 4.067 min
Delta R.T.: -0.003 min
Response: 158505
Conc: 3.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.668 min
Delta R.T.: -0.004 min
Response: 182322
Conc: 5.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.373 min
Delta R.T.: -0.009 min
Response: 207972
Conc: 6.21 ng/ml




#26 AR-1254-1

R.T.: 7.004 min
Delta R.T.: -0.003 min
Response: 71612
Conc: 55.75 ng/ml m

Instrument :
ECD_P
ClientSampleId :
72-11929-212


Manual Integrations
APPROVED

mohammad
11/23/2020 12:05:02 PM




#26 AR-1254-1

R.T.: 6.195 min
Delta R.T.: -0.006 min
Response: 79405
Conc: 41.73 ng/ml




#27 AR-1254-2

R.T.: 7.232 min
Delta R.T.: -0.004 min
Response: 97033
Conc: 50.26 ng/ml



#27 AR-1254-2

R.T.: 6.353 min
Delta R.T.: -0.007 min
Response: 51474
Conc: 31.45 ng/ml




#28 AR-1254-3

R.T.: 7.611 min
Delta R.T.: -0.004 min
Response: 131774
Conc: 60.79 ng/ml m

Instrument :
ECD_P
ClientSampleId :
72-11929-212


Manual Integrations
APPROVED

mohammad
11/23/2020 12:05:02 PM




#28 AR-1254-3

R.T.: 6.773 min
Delta R.T.: -0.008 min
Response: 86919
Conc: 31.84 ng/ml




#29 AR-1254-4

R.T.: 7.905 min
Delta R.T.: -0.004 min
Response: 42569
Conc: 28.19 ng/ml m



#29 AR-1254-4

R.T.: 7.013 min
Delta R.T.: -0.008 min
Response: 33880
Conc: 21.80 ng/ml




#30 AR-1254-5

R.T.: 8.330 min
Delta R.T.: -0.007 min
Response: 100299
Conc: 61.35 ng/ml m

Instrument :
ECD_P
ClientSampleId :
72-11929-212

Manual Integrations
APPROVED

mohammad
11/23/2020 12:05:02 PM



#30 AR-1254-5

R.T.: 7.440 min
Delta R.T.: -0.010 min
Response: 129759
Conc: 53.67 ng/ml