

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112020\
 Data File : PP031478.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Nov 2020 21:11
 Operator : DD\AJ
 Sample : L4863-04MS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S4-06-AMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 01:32:18 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 18 01:50:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.792	4.066	948446	838742	23.281	20.062
2) SA Decachlor...	10.666	9.372	490845	563960	15.981	16.842
Target Compounds						
3) L1 AR-1016-1	6.096	5.321	653068	605148	519.049	475.114
4) L1 AR-1016-2	6.119	5.342	989625	1140296	526.238	608.381
5) L1 AR-1016-3	6.186	5.533	634851	551217	554.631	551.671
6) L1 AR-1016-4	6.290	5.584	584496	373870	607.702	507.687
7) L1 AR-1016-5	6.605	5.814	526811	662276	601.056	681.327
31) L7 AR-1260-1	7.777	6.908	784408	829099	561.650	496.198
32) L7 AR-1260-2	8.038	7.105	1515227	1030357	910.567	506.306 #
33) L7 AR-1260-3	8.406	7.260	545138	913251	426.725	473.331
34) L7 AR-1260-4	8.635	7.743	667138	708406	436.397	439.272
35) L7 AR-1260-5	8.948	7.989	1176001	1525446	373.147	386.115


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112020\
Data File : PP031478.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Nov 2020 21:11
Operator : DD\AJ
Sample : L4863-04MS
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
S4-06-AMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 01:32:18 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111720.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 18 01:50:50 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.792 min
Delta R.T.: -0.004 min
Response: 948446
Conc: 23.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#1 Tetrachloro-m-xylene

R.T.: 4.066 min
Delta R.T.: -0.004 min
Response: 838742
Conc: 20.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: -0.006 min
Response: 490845
Conc: 15.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.372 min
Delta R.T.: -0.010 min
Response: 563960
Conc: 16.84 ng/ml



#3 AR-1016-1


R.T.: 6.096 min
Delta R.T.: -0.006 min
Response: 653068
Conc: 519.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#3 AR-1016-1

R.T.: 5.321 min
Delta R.T.: -0.009 min
Response: 605148
Conc: 475.11 ng/ml




#4 AR-1016-2

R.T.: 6.119 min
Delta R.T.: -0.006 min
Response: 989625
Conc: 526.24 ng/ml



#4 AR-1016-2


R.T.: 5.342 min
Delta R.T.: -0.008 min
Response: 1140296
Conc: 608.38 ng/ml



#5 AR-1016-3


R.T.: 6.186 min
Delta R.T.: -0.006 min
Response: 634851
Conc: 554.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#5 AR-1016-3

R.T.: 5.533 min
Delta R.T.: -0.010 min
Response: 551217
Conc: 551.67 ng/ml




#6 AR-1016-4

R.T.: 6.290 min
Delta R.T.: -0.007 min
Response: 584496
Conc: 607.70 ng/ml



#6 AR-1016-4


R.T.: 5.584 min
Delta R.T.: -0.010 min
Response: 373870
Conc: 507.69 ng/ml



#7 AR-1016-5


R.T.: 6.605 min
Delta R.T.: -0.006 min
Response: 526811
Conc: 601.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#7 AR-1016-5

R.T.: 5.814 min
Delta R.T.: -0.009 min
Response: 662276
Conc: 681.33 ng/ml




#31 AR-1260-1

R.T.: 7.777 min
Delta R.T.: -0.007 min
Response: 784408
Conc: 561.65 ng/ml



#31 AR-1260-1


R.T.: 6.908 min
Delta R.T.: -0.012 min
Response: 829099
Conc: 496.20 ng/ml



#32 AR-1260-2


R.T.: 8.038 min
Delta R.T.: -0.011 min
Response: 1515227
Conc: 910.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#32 AR-1260-2

R.T.: 7.105 min
Delta R.T.: -0.012 min
Response: 1030357
Conc: 506.31 ng/ml




#33 AR-1260-3

R.T.: 8.406 min
Delta R.T.: -0.008 min
Response: 545138
Conc: 426.73 ng/ml



#33 AR-1260-3


R.T.: 7.260 min
Delta R.T.: -0.012 min
Response: 913251
Conc: 473.33 ng/ml



#34 AR-1260-4


R.T.: 8.635 min
Delta R.T.: -0.009 min
Response: 667138
Conc: 436.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
S4-06-AMS




#34 AR-1260-4

R.T.: 7.743 min
Delta R.T.: -0.012 min
Response: 708406
Conc: 439.27 ng/ml



#35 AR-1260-5

R.T.: 8.948 min
Delta R.T.: -0.008 min
Response: 1176001
Conc: 373.15 ng/ml



#35 AR-1260-5

R.T.: 7.989 min
Delta R.T.: -0.012 min
Response: 1525446
Conc: 386.11 ng/ml