

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
 Data File : PP061770.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 15:03
 Operator : YP\AJ
 Sample : PB157262BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:22:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 04:03:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.727	27010763	35603701	20.497	16.923
2) SA Decachlor...	10.240	8.855	22984653	34364259	22.767	21.055
Target Compounds						
9) L2 AR-1221-2	4.761	4.034	526489	462535	43.346	23.407 #
27) L6 AR-1254-2	6.679f	0.000	6033089	0	69.791	N.D. #
28) L6 AR-1254-3	7.070	0.000	374990	0	4.409	N.D. #
32) L7 AR-1260-2	0.000	6.523f	0	512446	N.D.	4.427 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061770.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 15:03
Operator : YP\AJ
Sample : PB157262BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:22:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.002 min
Response: 27010763
Conc: 20.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.727 min
Delta R.T.: 0.000 min
Response: 35603701
Conc: 16.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.240 min
Delta R.T.: 0.005 min
Response: 22984653
Conc: 22.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.855 min
Delta R.T.: -0.005 min
Response: 34364259
Conc: 21.05 ng/ml



#9 AR-1221-2


R.T.: 4.761 min
Delta R.T.: 0.013 min
Response: 526489
Conc: 43.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 462535
Conc: 23.41 ng/ml




#27 AR-1254-2

R.T.: 6.679 min
Delta R.T.: -0.027 min
Response: 6033089
Conc: 69.79 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 5.812 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.070 min
Delta R.T.: -0.002 min
Response: 374990
Conc: 4.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.222 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.493 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 6.523 min
Delta R.T.: -0.021 min
Response: 512446
Conc: 4.43 ng/ml