

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
 Data File : PP061797.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 22:41
 Operator : YP\AJ
 Sample : PB157307BL
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:29:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 04:03:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.453	3.727	26544380	36285268	20.143	17.247
2) SA Decachlor...	10.239	8.853	22324019	33191624	22.113	20.336
Target Compounds						
5) L1 AR-1016-3	5.701	0.000	309286	0	6.844	N.D. #
9) L2 AR-1221-2	4.763	4.034	559625	529208	46.074	26.781 #
18) L4 AR-1242-3	5.701	0.000	309286	0	8.015	N.D. #
32) L7 AR-1260-2	0.000	6.523f	0	387485	N.D.	3.348 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061797.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 22:41
Operator : YP\AJ
Sample : PB157307BL
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:29:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: 0.003 min
Response: 26544380
Conc: 20.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.727 min
Delta R.T.: 0.000 min
Response: 36285268
Conc: 17.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.004 min
Response: 22324019
Conc: 22.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.853 min
Delta R.T.: -0.008 min
Response: 33191624
Conc: 20.34 ng/ml



#5 AR-1016-3


R.T.: 5.701 min
Delta R.T.: -0.009 min
Response: 309286
Conc: 6.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T.: 5.042 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.763 min
Delta R.T.: 0.015 min
Response: 559625
Conc: 46.07 ng/ml



#9 AR-1221-2


R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 529208
Conc: 26.78 ng/ml



#18 AR-1242-3


R.T.: 5.701 min
Delta R.T.: -0.010 min
Response: 309286
Conc: 8.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T. : 5.042 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.493 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 6.523 min
Delta R.T.: -0.021 min
Response: 387485
Conc: 3.35 ng/ml