

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
 Data File : PP061794.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 21:52
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:28:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 04:03:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.452	3.726	67326488	91480989	51.091	43.483
2)	SA Decachlor...	10.239	8.854	49221796	78548262	48.756	48.126
Target Compounds							
26)	L6 AR-1254-1	6.485	5.658	26996289	48002727	466.389	478.104
27)	L6 AR-1254-2	6.705	5.808	40124403	43329329	464.163	475.827
28)	L6 AR-1254-3	7.072	6.219	40200203	67343656	472.631	486.116
29)	L6 AR-1254-4	7.359	6.450	26488807	35939361	506.022	534.309
30)	L6 AR-1254-5	7.779	6.874	31060486	59425599	494.901	505.583

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061794.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 21:52
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:28:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

