

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 19:38
Operator : YP\AJ
Sample : 05460-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
W-A-3-CORONA

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:26:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.726	25583098	34971712	19.414	16.623
2) SA Decachlor...	10.239	8.853	18867333	28048723	18.689	17.185

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 19:38
Operator : YP\AJ
Sample : 05460-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
W-A-3-CORONA

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:26:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.002 min
Response: 25583098
Conc: 19.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
W-A-3-CORONA




#1 Tetrachloro-m-xylene

R.T.: 3.726 min
Delta R.T.: 0.000 min
Response: 34971712
Conc: 16.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.004 min
Response: 18867333
Conc: 18.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.853 min
Delta R.T.: -0.008 min
Response: 28048723
Conc: 17.19 ng/ml