

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
 Data File : PP061824.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 06:39
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 06:58:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 04:03:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.727	69128251	93014485	52.458	44.211
2) SA Decachlor...	10.236	8.852	49641381	80010539	49.171	49.022
Target Compounds						
26) L6 AR-1254-1	6.484	5.657	27488337	48719712	474.889	485.245
27) L6 AR-1254-2	6.704	5.806	41300719	43654373	477.771	479.396
28) L6 AR-1254-3	7.071	6.217	41147534	68139936	483.769	491.864
29) L6 AR-1254-4	7.358	6.448	26527198	35485437	506.755	527.561
30) L6 AR-1254-5	7.778	6.872	31222040	59714234	497.475	508.038



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112023\
Data File : PP061824.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 06:39
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 06:58:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.002 min
Response: 69128251
Conc: 52.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.727 min
Delta R.T.: 0.000 min
Response: 93014485
Conc: 44.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.001 min
Response: 49641381
Conc: 49.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.852 min
Delta R.T.: -0.009 min
Response: 80010539
Conc: 49.02 ng/ml



#26 AR-1254-1


R.T.: 6.484 min
Delta R.T.: -0.001 min
Response: 27488337
Conc: 474.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.006 min
Response: 48719712
Conc: 485.25 ng/ml




#27 AR-1254-2

R.T.: 6.704 min
Delta R.T.: -0.002 min
Response: 41300719
Conc: 477.77 ng/ml



#27 AR-1254-2


R.T.: 5.806 min
Delta R.T.: -0.006 min
Response: 43654373
Conc: 479.40 ng/ml



#28 AR-1254-3


R.T.: 7.071 min
Delta R.T.: -0.002 min
Response: 41147534
Conc: 483.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.217 min
Delta R.T.: -0.006 min
Response: 68139936
Conc: 491.86 ng/ml




#29 AR-1254-4

R.T.: 7.358 min
Delta R.T.: -0.002 min
Response: 26527198
Conc: 506.76 ng/ml



#29 AR-1254-4


R.T.: 6.448 min
Delta R.T.: -0.007 min
Response: 35485437
Conc: 527.56 ng/ml



#30 AR-1254-5

R.T.: 7.778 min
Delta R.T.: -0.003 min
Response: 31222040
Conc: 497.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.872 min
Delta R.T.: -0.007 min
Response: 59714234
Conc: 508.04 ng/ml