

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112122\
 Data File : PP053410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2022 13:55
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 17:35:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.644	92815667	83129020	48.283	55.263
2) SA Decachlor...	10.198	8.722	68790557	85779796	50.732	45.765
Target Compounds						
26) L6 AR-1254-1	6.432	5.557	31435132	43950241	444.649	484.831
27) L6 AR-1254-2	6.651	5.706	47109350	39325018	448.449	484.329
28) L6 AR-1254-3	7.019	6.115	46431984	63251141	452.671	497.592
29) L6 AR-1254-4	7.306	6.346	32220573	37964417	463.149	481.454
30) L6 AR-1254-5	7.726	6.768	36652573	58834243	462.745	465.794



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112122\
Data File : PP053410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2022 13:55
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 17:35:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.399 min
Delta R.T.: -0.004 min
Response: 92815667
Conc: 48.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.005 min
Response: 83129020
Conc: 55.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.198 min
Delta R.T.: -0.004 min
Response: 68790557
Conc: 50.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.722 min
Delta R.T.: -0.006 min
Response: 85779796
Conc: 45.76 ng/ml



#26 AR-1254-1


R.T.: 6.432 min
Delta R.T.: 0.000 min
Response: 31435132
Conc: 444.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.557 min
Delta R.T.: -0.003 min
Response: 43950241
Conc: 484.83 ng/ml




#27 AR-1254-2

R.T.: 6.651 min
Delta R.T.: -0.001 min
Response: 47109350
Conc: 448.45 ng/ml



#27 AR-1254-2


R.T.: 5.706 min
Delta R.T.: -0.003 min
Response: 39325018
Conc: 484.33 ng/ml



#28 AR-1254-3


R.T.: 7.019 min
Delta R.T.: -0.001 min
Response: 46431984
Conc: 452.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 63251141
Conc: 497.59 ng/ml




#29 AR-1254-4

R.T.: 7.306 min
Delta R.T.: 0.000 min
Response: 32220573
Conc: 463.15 ng/ml



#29 AR-1254-4


R.T.: 6.346 min
Delta R.T.: -0.002 min
Response: 37964417
Conc: 481.45 ng/ml



#30 AR-1254-5

R.T.: 7.726 min
Delta R.T.: -0.001 min
Response: 36652573
Conc: 462.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: -0.002 min
Response: 58834243
Conc: 465.79 ng/ml