

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112123\
 Data File : PP061894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2023 02:06
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/22/2023
 Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 02:28:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 16 04:03:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.454	3.725	69195096	97481755	52.509	46.335m
2) SA Decachlor...	10.247	8.853	53965243	95333336	53.454	58.410
Target Compounds						
3) L1 AR-1016-1	5.629	4.839	26684936	34405144	570.351	502.445
4) L1 AR-1016-2	5.652	4.859	40593805	48405497	569.427	511.646
5) L1 AR-1016-3	5.715	5.039	26341805	26796057	582.905	546.861
6) L1 AR-1016-4	5.813	5.081	20684530	22662364	566.879	522.470
7) L1 AR-1016-5	6.108	5.298	20386326	31674056	594.254	597.974
31) L7 AR-1260-1	7.240	6.350	36862788	58017023	531.351	576.722
32) L7 AR-1260-2	7.499	6.540	39362875	68031131	525.487	587.770
33) L7 AR-1260-3	7.859	6.697	27226767	65225034	532.556	595.115
34) L7 AR-1260-4	8.085	7.175	31707183	48678701	532.528	593.852
35) L7 AR-1260-5	8.408	7.418	52732386	107.3E6	533.582	621.868

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112123\
Data File : PP061894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2023 02:06
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 02:28:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 16 04:03:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

