

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112125\
 Data File : PP076606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2025 13:15
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 14:09:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 14:05:16 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.620	3.773	157.2E6	198.3E6	96.299	99.505
2) SA Decachlor...	10.375	8.771	125.1E6	156.2E6	95.027	97.910
Target Compounds						
16) L4 AR-1242-1	5.773	4.852	46820088	55846456	953.288	979.767
17) L4 AR-1242-2	5.794	4.870	69927193	82292663	942.657	975.914
18) L4 AR-1242-3	5.857	5.045	44603953	42952122	943.430	975.342
19) L4 AR-1242-4	5.954	5.130	36970364	41020216	944.933	964.969
20) L4 AR-1242-5	6.684	5.649	40380815	54892740	940.313	985.105



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112125\
Data File : PP076606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2025 13:15
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 14:09:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 14:05:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.620 min
Delta R.T.: 0.000 min
Response: 157180628
Conc: 96.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.773 min
Delta R.T.: 0.000 min
Response: 198256550
Conc: 99.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.375 min
Delta R.T.: 0.000 min
Response: 125095522
Conc: 95.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.771 min
Delta R.T.: 0.000 min
Response: 156160102
Conc: 97.91 ng/ml



#16 AR-1242-1


R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 46820088
Conc: 953.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 55846456
Conc: 979.77 ng/ml




#17 AR-1242-2

R.T.: 5.794 min
Delta R.T.: 0.000 min
Response: 69927193
Conc: 942.66 ng/ml



#17 AR-1242-2


R.T.: 4.870 min
Delta R.T.: 0.000 min
Response: 82292663
Conc: 975.91 ng/ml



#18 AR-1242-3


R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 44603953
Conc: 943.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 42952122
Conc: 975.34 ng/ml




#19 AR-1242-4

R.T.: 5.954 min
Delta R.T.: 0.000 min
Response: 36970364
Conc: 944.93 ng/ml



#19 AR-1242-4


R.T.: 5.130 min
Delta R.T.: 0.000 min
Response: 41020216
Conc: 964.97 ng/ml



#20 AR-1242-5

R.T.: 6.684 min
Delta R.T.: 0.000 min
Response: 40380815
Conc: 940.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.649 min
Delta R.T.: 0.000 min
Response: 54892740
Conc: 985.10 ng/ml