

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112220\
 Data File : PP031517.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Nov 2020 16:21
 Operator : DD\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 03:33:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 23 03:31:07 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.790	4.061	272465	254972	5.541	5.155
2)	SA Decachlor...	10.663	9.369	184302	210984	5.374	5.525
Target Compounds							
16)	L4 AR-1242-1	6.095	5.318	68495	61094	58.461	53.268
17)	L4 AR-1242-2	6.118	5.339	102224	89137	58.454	52.695
18)	L4 AR-1242-3	6.183	5.530	63497	47431	58.252	50.983
19)	L4 AR-1242-4	6.289	5.626	50311	44197	54.798	51.743
20)	L4 AR-1242-5	7.067	6.190	49418	52356	54.287	49.750



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112220\
Data File : PP031517.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Nov 2020 16:21
Operator : DD\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 03:33:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 23 03:31:07 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 272465
Conc: 5.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 4.061 min
Delta R.T.: 0.000 min
Response: 254972
Conc: 5.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.663 min
Delta R.T.: 0.000 min
Response: 184302
Conc: 5.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.369 min
Delta R.T.: 0.000 min
Response: 210984
Conc: 5.53 ng/ml



#16 AR-1242-1


R.T.: 6.095 min
Delta R.T.: 0.000 min
Response: 68495
Conc: 58.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 61094
Conc: 53.27 ng/ml




#17 AR-1242-2

R.T.: 6.118 min
Delta R.T.: 0.000 min
Response: 102224
Conc: 58.45 ng/ml



#17 AR-1242-2


R.T.: 5.339 min
Delta R.T.: 0.000 min
Response: 89137
Conc: 52.70 ng/ml



#18 AR-1242-3


R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 63497
Conc: 58.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 47431
Conc: 50.98 ng/ml




#19 AR-1242-4

R.T.: 6.289 min
Delta R.T.: 0.000 min
Response: 50311
Conc: 54.80 ng/ml



#19 AR-1242-4


R.T.: 5.626 min
Delta R.T.: 0.000 min
Response: 44197
Conc: 51.74 ng/ml



#20 AR-1242-5

R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 49418
Conc: 54.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 52356
Conc: 49.75 ng/ml