

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112221\
 Data File : PP041233.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Nov 2021 1:12
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 02:00:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.753	3.757	1078775	1462972	51.154	52.642
2)	SA Decachlor...	10.653	9.018	1126316	725776	51.562	50.629
Target Compounds							
3)	L1 AR-1016-1	6.069	5.005	368896	427765	524.696	504.048
4)	L1 AR-1016-2	6.093	5.023	565747	650338	509.288	529.170
5)	L1 AR-1016-3	6.158	5.214	358054	353841	503.265	525.131
6)	L1 AR-1016-4	6.265	5.267	287887	287706	510.375	529.036
7)	L1 AR-1016-5	6.578	5.494	282955	343796	511.655	516.025
8)	L2 AR-1221-1	4.996	4.013	39700	52981	175.239	165.954
9)	L2 AR-1221-2	5.102	4.112	49644	75795	320.692	298.922
10)	L2 AR-1221-3	5.189	4.200	198940	301631	371.886	401.455
11)	L3 AR-1232-1	5.189	4.200	198940	301631	449.054	503.635
12)	L3 AR-1232-2	5.773	5.023	289389	650338	1253.993	1308.340
13)	L3 AR-1232-3	6.093	5.214	565747	353841	1323.459	1321.538
14)	L3 AR-1232-4	6.265	5.310	287887	344590	1416.659	1468.860
15)	L3 AR-1232-5	6.362	5.494	226601	343796	1605.191	1398.052
16)	L4 AR-1242-1	6.069	5.005	368896	427765	713.641	673.923
17)	L4 AR-1242-2	6.093	5.023	565747	650338	692.592	707.184
18)	L4 AR-1242-3	6.158	5.214	358054	353841	681.317	697.029
19)	L4 AR-1242-4	6.265	5.310	287887	344590	713.372	701.977
20)	L4 AR-1242-5	7.046	5.873	43290	246299	96.840	456.187 #
21)	L5 AR-1248-1	6.069	5.005	368896	427765	992.872	885.502
22)	L5 AR-1248-2	6.362	5.267	226601	287706	408.656	424.067
23)	L5 AR-1248-3	6.578	5.310	282955	344590	420.202	493.840
24)	L5 AR-1248-4	7.006	5.494	53691	343796	71.964	429.400 #
25)	L5 AR-1248-5	7.046	5.915	43290	42721	57.256	59.010
26)	L6 AR-1254-1	6.976	5.873	189382	246299	245.228	216.296
27)	L6 AR-1254-2	7.204	6.032	207339	219072	167.744	223.420 #
28)	L6 AR-1254-3	7.586	6.469	129969	408556	97.061	282.450 #
29)	L6 AR-1254-4	7.880	6.692	92753	58983	97.192	71.107 #
30)	L6 AR-1254-5	8.307	7.120	695595	623353	654.421	534.766
31)	L7 AR-1260-1	7.750	6.588	513371	530945	520.840	554.606
32)	L7 AR-1260-2	8.017	6.787	591049	602841	500.131	555.612
33)	L7 AR-1260-3	8.381	6.940	460031	552015	506.748	550.157
34)	L7 AR-1260-4	8.610	7.422	543268	440595	502.791	537.190
35)	L7 AR-1260-5	8.925	7.673	1035952	925705	497.094	530.406
36)	L8 AR-1262-1	8.381	7.120	460031	623353	385.024	1052.339 #
37)	L8 AR-1262-2	8.925	7.673	1035952	925705	491.149	559.258
38)	L8 AR-1262-3	9.222	7.958	237435	230689	235.217	320.261 #
39)	L8 AR-1262-4	9.310	8.021	208007	694845	311.653	540.375 #
40)	L8 AR-1262-5	9.945	8.522	348575	259081	423.193	431.493
41)	L9 AR-1268-1	9.222	7.958	237435	230689	87.264	110.752 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112221\
 Data File : PP041233.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Nov 2021 1:12
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 02:00:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.310	8.021	208007	694845	80.838	364.662 #
43) L9	AR-1268-3	9.526	8.228	15772	13481	7.203	8.301
44) L9	AR-1268-4	9.945	8.522	348575	259081	363.263	368.568
45) L9	AR-1268-5	10.337	8.793	105007	71480	14.861	15.384



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112221\
Data File : PP041233.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Nov 2021 1:12
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 02:00:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.753 min
Delta R.T.: -0.004 min
Response: 1078775
Conc: 51.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.757 min
Delta R.T.: -0.003 min
Response: 1462972
Conc: 52.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.653 min
Delta R.T.: -0.008 min
Response: 1126316
Conc: 51.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.018 min
Delta R.T.: -0.008 min
Response: 725776
Conc: 50.63 ng/ml



#3 AR-1016-1


R.T.: 6.069 min
Delta R.T.: -0.006 min
Response: 368896
Conc: 524.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 427765
Conc: 504.05 ng/ml




#4 AR-1016-2

R.T.: 6.093 min
Delta R.T.: -0.006 min
Response: 565747
Conc: 509.29 ng/ml



#4 AR-1016-2


R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 650338
Conc: 529.17 ng/ml



#5 AR-1016-3


R.T.: 6.158 min
Delta R.T.: -0.006 min
Response: 358054
Conc: 503.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.214 min
Delta R.T.: -0.005 min
Response: 353841
Conc: 525.13 ng/ml




#6 AR-1016-4

R.T.: 6.265 min
Delta R.T.: -0.005 min
Response: 287887
Conc: 510.37 ng/ml



#6 AR-1016-4


R.T.: 5.267 min
Delta R.T.: -0.005 min
Response: 287706
Conc: 529.04 ng/ml



#7 AR-1016-5


R.T.: 6.578 min
Delta R.T.: -0.005 min
Response: 282955
Conc: 511.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.494 min
Delta R.T.: -0.005 min
Response: 343796
Conc: 516.02 ng/ml




#8 AR-1221-1

R.T.: 4.996 min
Delta R.T.: -0.014 min
Response: 39700
Conc: 175.24 ng/ml



#8 AR-1221-1


R.T.: 4.013 min
Delta R.T.: -0.004 min
Response: 52981
Conc: 165.95 ng/ml



#9 AR-1221-2


R.T.: 5.102 min
Delta R.T.: -0.005 min
Response: 49644
Conc: 320.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.112 min
Delta R.T.: -0.004 min
Response: 75795
Conc: 298.92 ng/ml




#10 AR-1221-3

R.T.: 5.189 min
Delta R.T.: -0.005 min
Response: 198940
Conc: 371.89 ng/ml



#10 AR-1221-3


R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 301631
Conc: 401.46 ng/ml



#11 AR-1232-1


R.T.: 5.189 min
Delta R.T.: -0.005 min
Response: 198940
Conc: 449.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 301631
Conc: 503.64 ng/ml




#12 AR-1232-2

R.T.: 5.773 min
Delta R.T.: -0.004 min
Response: 289389
Conc: 1253.99 ng/ml



#12 AR-1232-2


R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 650338
Conc: 1308.34 ng/ml



#13 AR-1232-3


R.T.: 6.093 min
Delta R.T.: -0.005 min
Response: 565747
Conc: 1323.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.214 min
Delta R.T.: -0.004 min
Response: 353841
Conc: 1321.54 ng/ml




#14 AR-1232-4

R.T.: 6.265 min
Delta R.T.: -0.005 min
Response: 287887
Conc: 1416.66 ng/ml



#14 AR-1232-4


R.T.: 5.310 min
Delta R.T.: -0.005 min
Response: 344590
Conc: 1468.86 ng/ml



#15 AR-1232-5


R.T.: 6.362 min
Delta R.T.: -0.005 min
Response: 226601
Conc: 1605.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.494 min
Delta R.T.: -0.004 min
Response: 343796
Conc: 1398.05 ng/ml




#16 AR-1242-1

R.T.: 6.069 min
Delta R.T.: -0.005 min
Response: 368896
Conc: 713.64 ng/ml



#16 AR-1242-1


R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 427765
Conc: 673.92 ng/ml



#17 AR-1242-2


R.T.: 6.093 min
Delta R.T.: -0.005 min
Response: 565747
Conc: 692.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 650338
Conc: 707.18 ng/ml




#18 AR-1242-3

R.T.: 6.158 min
Delta R.T.: -0.005 min
Response: 358054
Conc: 681.32 ng/ml



#18 AR-1242-3


R.T.: 5.214 min
Delta R.T.: -0.004 min
Response: 353841
Conc: 697.03 ng/ml



#19 AR-1242-4


R.T.: 6.265 min
Delta R.T.: -0.005 min
Response: 287887
Conc: 713.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.310 min
Delta R.T.: -0.004 min
Response: 344590
Conc: 701.98 ng/ml




#20 AR-1242-5

R.T.: 7.046 min
Delta R.T.: -0.005 min
Response: 43290
Conc: 96.84 ng/ml



#20 AR-1242-5


R.T.: 5.873 min
Delta R.T.: -0.006 min
Response: 246299
Conc: 456.19 ng/ml



#21 AR-1248-1


R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 368896
Conc: 992.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.005 min
Delta R.T.: -0.004 min
Response: 427765
Conc: 885.50 ng/ml




#22 AR-1248-2

R.T.: 6.362 min
Delta R.T.: -0.004 min
Response: 226601
Conc: 408.66 ng/ml



#22 AR-1248-2


R.T.: 5.267 min
Delta R.T.: -0.004 min
Response: 287706
Conc: 424.07 ng/ml



#23 AR-1248-3


R.T.: 6.578 min
Delta R.T.: -0.004 min
Response: 282955
Conc: 420.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.310 min
Delta R.T.: -0.004 min
Response: 344590
Conc: 493.84 ng/ml




#24 AR-1248-4

R.T.: 7.006 min
Delta R.T.: -0.005 min
Response: 53691
Conc: 71.96 ng/ml



#24 AR-1248-4


R.T.: 5.494 min
Delta R.T.: -0.004 min
Response: 343796
Conc: 429.40 ng/ml



#25 AR-1248-5


R.T.: 7.046 min
Delta R.T.: -0.005 min
Response: 43290
Conc: 57.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 5.915 min
Delta R.T.: -0.005 min
Response: 42721
Conc: 59.01 ng/ml




#26 AR-1254-1

R.T.: 6.976 min
Delta R.T.: -0.003 min
Response: 189382
Conc: 245.23 ng/ml



#26 AR-1254-1


R.T.: 5.873 min
Delta R.T.: -0.004 min
Response: 246299
Conc: 216.30 ng/ml



#27 AR-1254-2


R.T.: 7.204 min
Delta R.T.: -0.004 min
Response: 207339
Conc: 167.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.032 min
Delta R.T.: -0.005 min
Response: 219072
Conc: 223.42 ng/ml




#28 AR-1254-3

R.T.: 7.586 min
Delta R.T.: -0.003 min
Response: 129969
Conc: 97.06 ng/ml



#28 AR-1254-3


R.T.: 6.469 min
Delta R.T.: 0.013 min
Response: 408556
Conc: 282.45 ng/ml



#29 AR-1254-4


R.T.: 7.880 min
Delta R.T.: -0.005 min
Response: 92753
Conc: 97.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 6.692 min
Delta R.T.: -0.005 min
Response: 58983
Conc: 71.11 ng/ml




#30 AR-1254-5

R.T.: 8.307 min
Delta R.T.: -0.005 min
Response: 695595
Conc: 654.42 ng/ml



#30 AR-1254-5


R.T.: 7.120 min
Delta R.T.: -0.005 min
Response: 623353
Conc: 534.77 ng/ml



#31 AR-1260-1


R.T.: 7.750 min
Delta R.T.: -0.006 min
Response: 513371
Conc: 520.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.588 min
Delta R.T.: -0.006 min
Response: 530945
Conc: 554.61 ng/ml




#32 AR-1260-2

R.T.: 8.017 min
Delta R.T.: -0.005 min
Response: 591049
Conc: 500.13 ng/ml



#32 AR-1260-2


R.T.: 6.787 min
Delta R.T.: -0.006 min
Response: 602841
Conc: 555.61 ng/ml



#33 AR-1260-3


R.T.: 8.381 min
Delta R.T.: -0.005 min
Response: 460031
Conc: 506.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 6.940 min
Delta R.T.: -0.006 min
Response: 552015
Conc: 550.16 ng/ml




#34 AR-1260-4

R.T.: 8.610 min
Delta R.T.: -0.006 min
Response: 543268
Conc: 502.79 ng/ml



#34 AR-1260-4


R.T.: 7.422 min
Delta R.T.: -0.006 min
Response: 440595
Conc: 537.19 ng/ml



#35 AR-1260-5


R.T.: 8.925 min
Delta R.T.: -0.006 min
Response: 1035952
Conc: 497.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.673 min
Delta R.T.: -0.006 min
Response: 925705
Conc: 530.41 ng/ml




#36 AR-1262-1

R.T.: 8.381 min
Delta R.T.: -0.003 min
Response: 460031
Conc: 385.02 ng/ml



#36 AR-1262-1


R.T.: 7.120 min
Delta R.T.: -0.003 min
Response: 623353
Conc: 1052.34 ng/ml



#37 AR-1262-2


R.T.: 8.925 min
Delta R.T.: -0.005 min
Response: 1035952
Conc: 491.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.673 min
Delta R.T.: -0.004 min
Response: 925705
Conc: 559.26 ng/ml




#38 AR-1262-3

R.T.: 9.222 min
Delta R.T.: -0.004 min
Response: 237435
Conc: 235.22 ng/ml



#38 AR-1262-3


R.T.: 7.958 min
Delta R.T.: -0.004 min
Response: 230689
Conc: 320.26 ng/ml



#39 AR-1262-4


R.T.: 9.310 min
Delta R.T.: -0.005 min
Response: 208007
Conc: 311.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.021 min
Delta R.T.: -0.005 min
Response: 694845
Conc: 540.38 ng/ml




#40 AR-1262-5

R.T.: 9.945 min
Delta R.T.: -0.006 min
Response: 348575
Conc: 423.19 ng/ml



#40 AR-1262-5


R.T.: 8.522 min
Delta R.T.: -0.004 min
Response: 259081
Conc: 431.49 ng/ml



#41 AR-1268-1


R.T.: 9.222 min
Delta R.T.: -0.001 min
Response: 237435
Conc: 87.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 7.958 min
Delta R.T.: -0.005 min
Response: 230689
Conc: 110.75 ng/ml




#42 AR-1268-2

R.T.: 9.310 min
Delta R.T.: -0.006 min
Response: 208007
Conc: 80.84 ng/ml



#42 AR-1268-2


R.T.: 8.021 min
Delta R.T.: -0.006 min
Response: 694845
Conc: 364.66 ng/ml



#43 AR-1268-3


R.T.: 9.526 min
Delta R.T.: -0.004 min
Response: 15772
Conc: 7.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 8.228 min
Delta R.T.: -0.004 min
Response: 13481
Conc: 8.30 ng/ml




#44 AR-1268-4

R.T.: 9.945 min
Delta R.T.: -0.004 min
Response: 348575
Conc: 363.26 ng/ml



#44 AR-1268-4


R.T.: 8.522 min
Delta R.T.: -0.004 min
Response: 259081
Conc: 368.57 ng/ml



#45 AR-1268-5

R.T.: 10.337 min
Delta R.T.: -0.003 min
Response: 105007
Conc: 14.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 8.793 min
Delta R.T.: -0.005 min
Response: 71480
Conc: 15.38 ng/ml