

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112320\
 Data File : PP031564.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Nov 2020 15:45
 Operator : DD\AJ
 Sample : L4850-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WOOSPK-038-0002

Manual Integrations
 APPROVED

Ankita
 11/25/2020 4:35:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 01:24:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 23 04:59:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.790	4.063	697861	750271	14.412	15.657m
2) SA Decachlor...	10.667	9.371	724725	685988	21.338	18.527

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112320\
Data File : PP031564.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Nov 2020 15:45
Operator : DD\AJ
Sample : L4850-01
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WOOSPK-038-0002


Manual Integrations
APPROVED

Ankita
11/25/2020 4:35:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 01:24:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 23 04:59:03 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 697861
Conc: 14.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
WOOSPK-038-0002


Manual Integrations
APPROVED

Ankita
11/25/2020 4:35:27 PM




#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: 0.000 min
Response: 750271
Conc: 15.66 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.667 min
Delta R.T.: 0.001 min
Response: 724725
Conc: 21.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.371 min
Delta R.T.: 0.000 min
Response: 685988
Conc: 18.53 ng/ml