

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112320\
 Data File : PP031598.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Nov 2020 1:46
 Operator : DD\AJ
 Sample : L4821-08
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WOOSPK-040-0002

Manual Integrations
 APPROVED

Ankita
 11/25/2020 5:28:54 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 04:31:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 23 04:59:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.789	4.063	861493	877543	17.791m	18.313
2) SA Decachlor...	10.663	9.369	701633	359581	20.658	9.711m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112320\
Data File : PP031598.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Nov 2020 1:46
Operator : DD\AJ
Sample : L4821-08
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WOOSPK-040-0002


Manual Integrations
APPROVED

Ankita
11/25/2020 5:28:54 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 04:31:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 23 04:59:03 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 861493
Conc: 17.79 ng/ml m

Instrument :
ECD_P
ClientSampleId :
WOOSPK-040-0002


Manual Integrations
APPROVED

Ankita
11/25/2020 5:28:54 PM




#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: 0.001 min
Response: 877543
Conc: 18.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.663 min
Delta R.T.: -0.002 min
Response: 701633
Conc: 20.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.369 min
Delta R.T.: -0.002 min
Response: 359581
Conc: 9.71 ng/ml m