

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
 Data File : PP031668.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Nov 2020 10:19
 Operator : DD\AJ
 Sample : L4822-09DL 10X
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 FD-02DL

Manual Integrations
 APPROVED

Ankita
 12/2/2020 2:28:50 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 12:01:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.788	4.061	90734	77972	1.874	1.627
2) SA Decachlor...	10.658	9.366	86102	140937	2.535	3.806m#
Target Compounds						
36) L8 AR-1262-1	8.403	7.474	408903	482058	179.970	173.450
37) L8 AR-1262-2	8.945	7.984	738767	872356	185.919	179.595
38) L8 AR-1262-3	9.243	8.269	599033	538155	215.382	273.606 #
39) L8 AR-1262-4	9.333	8.335	717927	963947	327.502	265.903
40) L8 AR-1262-5	9.957	8.830	473044	445026	323.998	274.168

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
Data File : PP031668.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Nov 2020 10:19
Operator : DD\AJ
Sample : L4822-09DL 10X
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
FD-02DL

Manual Integrations
APPROVED

Ankita
12/2/2020 2:28:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 12:01:24 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

