

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
 Data File : PP031620.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Nov 2020 14:17
 Operator : DD\AJ
 Sample : L4850-18
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 2020-11-19-EB

Manual Integrations
 APPROVED

Ankita
 11/25/2020 5:39:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 00:39:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.792	4.064	904065	868213	18.670	18.119m
2) SA Decachlor...	10.666	9.367	828384	374866	24.390m	10.124m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
Data File : PP031620.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Nov 2020 14:17
Operator : DD\AJ
Sample : L4850-18
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
2020-11-19-EB


Manual Integrations
APPROVED

Ankita
11/25/2020 5:39:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 00:39:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.792 min
Delta R.T.: 0.002 min
Response: 904065
Conc: 18.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
2020-11-19-EB


Manual Integrations
APPROVED

Ankita
11/25/2020 5:39:58 PM




#1 Tetrachloro-m-xylene

R.T.: 4.064 min
Delta R.T.: 0.003 min
Response: 868213
Conc: 18.12 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: 0.000 min
Response: 828384
Conc: 24.39 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.367 min
Delta R.T.: -0.004 min
Response: 374866
Conc: 10.12 ng/ml m