

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
 Data File : PP031626.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Nov 2020 15:57
 Operator : DD\AJ
 Sample : L4868-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 KENS-B3-112020-8-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 00:43:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.791	4.063	1111842	1064337	22.961	22.211
2) SA Decachlor...	10.663	9.366	688712	699683	20.278	18.896

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112420\
Data File : PP031626.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Nov 2020 15:57
Operator : DD\AJ
Sample : L4868-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
KENS-B3-112020-8-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 00:43:00 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 1111842
Conc: 22.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
KENS-B3-112020-8-10




#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: 0.001 min
Response: 1064337
Conc: 22.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.663 min
Delta R.T.: -0.002 min
Response: 688712
Conc: 20.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.366 min
Delta R.T.: -0.004 min
Response: 699683
Conc: 18.90 ng/ml