

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112524\
 Data File : PP068727.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2024 06:44
 Operator : YP\AJ
 Sample : P4999-01 10X
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 Y2309-0009-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/26/2024
 Supervised By :Ankita Jodhani 11/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 10:19:48 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 11:04:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.761	4.050	1327002	1392824	1.367	1.438
2) SA Decachlor...	10.677	9.206	2217870	1682664	1.772m	1.488
Target Compounds						
26) L6 AR-1254-1	6.780	5.970	5400026	6303053	104.074m	103.939
27) L6 AR-1254-2	6.998	6.117	6883057	5085918	100.869	96.442
28) L6 AR-1254-3	7.363	6.526	5623138	6800016	78.691	81.671
29) L6 AR-1254-4	7.646	6.754	3091401	2836724	57.446	58.336
30) L6 AR-1254-5	8.063	7.175	3244144	3754417	51.768m	50.305

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112524\
Data File : PP068727.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2024 06:44
Operator : YP\AJ
Sample : P4999-01 10X
Misc :
ALS Vial : 40 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
Y2309-0009-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 10:19:48 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110724.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 11:04:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.761 min
Delta R.T.: 0.004 min
Response: 1327002
Conc: 1.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
Y2309-0009-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024




#1 Tetrachloro-m-xylene

R.T.: 4.050 min
Delta R.T.: -0.001 min
Response: 1392824
Conc: 1.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.677 min
Delta R.T.: 0.005 min
Response: 2217870
Conc: 1.77 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.206 min
Delta R.T.: -0.004 min
Response: 1682664
Conc: 1.49 ng/ml




#26 AR-1254-1

R.T.: 6.780 min
Delta R.T.: 0.006 min
Response: 5400026
Conc: 104.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
Y2309-0009-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024




#26 AR-1254-1

R.T.: 5.970 min
Delta R.T.: -0.002 min
Response: 6303053
Conc: 103.94 ng/ml




#27 AR-1254-2

R.T.: 6.998 min
Delta R.T.: 0.006 min
Response: 6883057
Conc: 100.87 ng/ml



#27 AR-1254-2

R.T.: 6.117 min
Delta R.T.: -0.001 min
Response: 5085918
Conc: 96.44 ng/ml




#28 AR-1254-3

R.T.: 7.363 min
Delta R.T.: 0.006 min
Response: 5623138
Conc: 78.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
Y2309-0009-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024




#28 AR-1254-3

R.T.: 6.526 min
Delta R.T.: -0.002 min
Response: 6800016
Conc: 81.67 ng/ml




#29 AR-1254-4

R.T.: 7.646 min
Delta R.T.: 0.006 min
Response: 3091401
Conc: 57.45 ng/ml



#29 AR-1254-4

R.T.: 6.754 min
Delta R.T.: -0.002 min
Response: 2836724
Conc: 58.34 ng/ml




#30 AR-1254-5

R.T.: 8.063 min
Delta R.T.: 0.006 min
Response: 3244144
Conc: 51.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
Y2309-0009-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024



#30 AR-1254-5

R.T.: 7.175 min
Delta R.T.: -0.002 min
Response: 3754417
Conc: 50.30 ng/ml