

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112822\
 Data File : PP053515.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Nov 2022 12:03
 Operator : YP\AJ
 Sample : N5771-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PCB-112222-AP-04

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 16:07:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.402	3.648	40616704	36787735	21.129	24.456
2) SA Decachlor...	10.210	8.729	32741241	37888170	24.146	20.214

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112822\
Data File : PP053515.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Nov 2022 12:03
Operator : YP\AJ
Sample : N5771-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PCB-112222-AP-04

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 16:07:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 40616704
Conc: 21.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
PCB-112222-AP-04




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: -0.001 min
Response: 36787735
Conc: 24.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: 0.008 min
Response: 32741241
Conc: 24.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 37888170
Conc: 20.21 ng/ml