

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
 Data File : PP041391.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Nov 2021 14:58
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 00:47:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.751	3.756	1124010	1416863	53.299	50.983
2)	SA Decachlor...	10.648	9.015	1048023	702662	47.978	49.016
Target Compounds							
3)	L1 AR-1016-1	6.067	5.003	365665	415133	520.100	489.165
4)	L1 AR-1016-2	6.090	5.021	567709	642013	511.054	522.397
5)	L1 AR-1016-3	6.155	5.212	354085	347212	497.686	515.293
6)	L1 AR-1016-4	6.262	5.264	283549	280523	502.684	515.829
7)	L1 AR-1016-5	6.575	5.492	282129	331481	510.161	497.542
8)	L2 AR-1221-1	4.993f	4.011	39583	47943	174.723	150.177
9)	L2 AR-1221-2	5.100	4.110	50995	74721	329.422	294.684
10)	L2 AR-1221-3	5.186	4.198	204608	294409	382.482	391.843
11)	L3 AR-1232-1	5.186	4.198	204608	294409	461.849	491.576
12)	L3 AR-1232-2	5.770	5.021	295818	642013	1281.854	1291.592
13)	L3 AR-1232-3	6.090	5.212	567709	347212	1328.048	1296.781
14)	L3 AR-1232-4	6.262	5.308	283549	330977	1395.312	1410.833
15)	L3 AR-1232-5	6.359	5.492	224335	331481	1589.142	1347.976
16)	L4 AR-1242-1	6.067	5.003	365665	415133	707.390	654.022
17)	L4 AR-1242-2	6.090	5.021	567709	642013	694.993	698.132
18)	L4 AR-1242-3	6.155	5.212	354085	347212	673.765	683.971
19)	L4 AR-1242-4	6.262	5.308	283549	330977	702.622	674.246
20)	L4 AR-1242-5	7.043	5.870	48215	238672	107.859	442.059 #
21)	L5 AR-1248-1	6.067	5.003	365665	415133	984.175	859.354
22)	L5 AR-1248-2	6.359	5.264	224335	280523	404.570	413.481
23)	L5 AR-1248-3	6.575	5.308	282129	330977	418.975	474.331
24)	L5 AR-1248-4	7.004	5.492	59342	331481	79.539	414.019 #
25)	L5 AR-1248-5	7.043	5.914	48215	42041	63.770	58.071
26)	L6 AR-1254-1	6.972	5.870	189667	238672	245.598	209.597
27)	L6 AR-1254-2	7.200	6.030	209407	213706	169.418	217.948 #
28)	L6 AR-1254-3	7.582	6.467	122540	396635	91.513	274.209 #
29)	L6 AR-1254-4	7.877	6.690	86906	56377	91.065	67.965 #
30)	L6 AR-1254-5	8.304	7.118	677487	607381	637.385	521.064
31)	L7 AR-1260-1	7.747	6.586	501525	517763	508.822	540.837
32)	L7 AR-1260-2	8.014	6.785	575541	585976	487.008	540.068
33)	L7 AR-1260-3	8.377	6.938	446836	539127	492.213	537.313
34)	L7 AR-1260-4	8.608	7.420	524515	429149	485.435	523.235
35)	L7 AR-1260-5	8.923	7.671	997331	909265	478.562	520.986
36)	L8 AR-1262-1	8.377	7.118	446836	607381	373.980	1025.375 #
37)	L8 AR-1262-2	8.923	7.671	997331	909265	472.838	549.326
38)	L8 AR-1262-3	9.221	7.956	238026	227373	235.803	315.657 #
39)	L8 AR-1262-4	9.289f	8.019	457746	688678	685.832	535.579
40)	L8 AR-1262-5	9.941	8.519	322943	249279	392.074	415.169
41)	L9 AR-1268-1	9.221	7.956	238026	227373	87.482	109.160

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
 Data File : PP041391.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Nov 2021 14:58
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 00:47:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.289f	8.019	457746	688678	177.895	361.425 #
43) L9	AR-1268-3	9.520	8.224	13466	12105	6.150	7.454
44) L9	AR-1268-4	9.941	8.519	322943	249279	336.550	354.625
45) L9	AR-1268-5	10.333	8.790	93783	70717	13.273	15.220



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
Data File : PP041391.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Nov 2021 14:58
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 00:47:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.751 min
Delta R.T.: -0.006 min
Response: 1124010
Conc: 53.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.756 min
Delta R.T.: -0.004 min
Response: 1416863
Conc: 50.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.648 min
Delta R.T.: -0.013 min
Response: 1048023
Conc: 47.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.015 min
Delta R.T.: -0.010 min
Response: 702662
Conc: 49.02 ng/ml



#3 AR-1016-1


R.T.: 6.067 min
Delta R.T.: -0.008 min
Response: 365665
Conc: 520.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.003 min
Delta R.T.: -0.006 min
Response: 415133
Conc: 489.16 ng/ml




#4 AR-1016-2

R.T.: 6.090 min
Delta R.T.: -0.009 min
Response: 567709
Conc: 511.05 ng/ml



#4 AR-1016-2


R.T.: 5.021 min
Delta R.T.: -0.007 min
Response: 642013
Conc: 522.40 ng/ml



#5 AR-1016-3


R.T.: 6.155 min
Delta R.T.: -0.008 min
Response: 354085
Conc: 497.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.212 min
Delta R.T.: -0.007 min
Response: 347212
Conc: 515.29 ng/ml




#6 AR-1016-4

R.T.: 6.262 min
Delta R.T.: -0.008 min
Response: 283549
Conc: 502.68 ng/ml



#6 AR-1016-4


R.T.: 5.264 min
Delta R.T.: -0.007 min
Response: 280523
Conc: 515.83 ng/ml



#7 AR-1016-5


R.T.: 6.575 min
Delta R.T.: -0.008 min
Response: 282129
Conc: 510.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.492 min
Delta R.T.: -0.008 min
Response: 331481
Conc: 497.54 ng/ml




#8 AR-1221-1

R.T.: 4.993 min
Delta R.T.: -0.017 min
Response: 39583
Conc: 174.72 ng/ml



#8 AR-1221-1


R.T.: 4.011 min
Delta R.T.: -0.006 min
Response: 47943
Conc: 150.18 ng/ml



#9 AR-1221-2


R.T.: 5.100 min
Delta R.T.: -0.008 min
Response: 50995
Conc: 329.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.110 min
Delta R.T.: -0.006 min
Response: 74721
Conc: 294.68 ng/ml




#10 AR-1221-3

R.T.: 5.186 min
Delta R.T.: -0.008 min
Response: 204608
Conc: 382.48 ng/ml



#10 AR-1221-3


R.T.: 4.198 min
Delta R.T.: -0.005 min
Response: 294409
Conc: 391.84 ng/ml



#11 AR-1232-1


R.T.: 5.186 min
Delta R.T.: -0.008 min
Response: 204608
Conc: 461.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.198 min
Delta R.T.: -0.005 min
Response: 294409
Conc: 491.58 ng/ml




#12 AR-1232-2

R.T.: 5.770 min
Delta R.T.: -0.007 min
Response: 295818
Conc: 1281.85 ng/ml



#12 AR-1232-2


R.T.: 5.021 min
Delta R.T.: -0.006 min
Response: 642013
Conc: 1291.59 ng/ml



#13 AR-1232-3


R.T.: 6.090 min
Delta R.T.: -0.008 min
Response: 567709
Conc: 1328.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.212 min
Delta R.T.: -0.006 min
Response: 347212
Conc: 1296.78 ng/ml




#14 AR-1232-4

R.T.: 6.262 min
Delta R.T.: -0.008 min
Response: 283549
Conc: 1395.31 ng/ml



#14 AR-1232-4


R.T.: 5.308 min
Delta R.T.: -0.007 min
Response: 330977
Conc: 1410.83 ng/ml



#15 AR-1232-5


R.T.: 6.359 min
Delta R.T.: -0.008 min
Response: 224335
Conc: 1589.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.492 min
Delta R.T.: -0.006 min
Response: 331481
Conc: 1347.98 ng/ml




#16 AR-1242-1

R.T.: 6.067 min
Delta R.T.: -0.008 min
Response: 365665
Conc: 707.39 ng/ml



#16 AR-1242-1


R.T.: 5.003 min
Delta R.T.: -0.006 min
Response: 415133
Conc: 654.02 ng/ml



#17 AR-1242-2


R.T.: 6.090 min
Delta R.T.: -0.008 min
Response: 567709
Conc: 694.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.021 min
Delta R.T.: -0.006 min
Response: 642013
Conc: 698.13 ng/ml




#18 AR-1242-3

R.T.: 6.155 min
Delta R.T.: -0.007 min
Response: 354085
Conc: 673.77 ng/ml



#18 AR-1242-3


R.T.: 5.212 min
Delta R.T.: -0.006 min
Response: 347212
Conc: 683.97 ng/ml



#19 AR-1242-4


R.T.: 6.262 min
Delta R.T.: -0.008 min
Response: 283549
Conc: 702.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.308 min
Delta R.T.: -0.006 min
Response: 330977
Conc: 674.25 ng/ml




#20 AR-1242-5

R.T.: 7.043 min
Delta R.T.: -0.008 min
Response: 48215
Conc: 107.86 ng/ml



#20 AR-1242-5


R.T.: 5.870 min
Delta R.T.: -0.008 min
Response: 238672
Conc: 442.06 ng/ml



#21 AR-1248-1


R.T.: 6.067 min
Delta R.T.: -0.007 min
Response: 365665
Conc: 984.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.003 min
Delta R.T.: -0.006 min
Response: 415133
Conc: 859.35 ng/ml




#22 AR-1248-2

R.T.: 6.359 min
Delta R.T.: -0.007 min
Response: 224335
Conc: 404.57 ng/ml



#22 AR-1248-2


R.T.: 5.264 min
Delta R.T.: -0.006 min
Response: 280523
Conc: 413.48 ng/ml



#23 AR-1248-3


R.T.: 6.575 min
Delta R.T.: -0.007 min
Response: 282129
Conc: 418.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.308 min
Delta R.T.: -0.006 min
Response: 330977
Conc: 474.33 ng/ml




#24 AR-1248-4

R.T.: 7.004 min
Delta R.T.: -0.008 min
Response: 59342
Conc: 79.54 ng/ml



#24 AR-1248-4


R.T.: 5.492 min
Delta R.T.: -0.007 min
Response: 331481
Conc: 414.02 ng/ml



#25 AR-1248-5


R.T.: 7.043 min
Delta R.T.: -0.007 min
Response: 48215
Conc: 63.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.914 min
Delta R.T.: -0.007 min
Response: 42041
Conc: 58.07 ng/ml




#26 AR-1254-1

R.T.: 6.972 min
Delta R.T.: -0.007 min
Response: 189667
Conc: 245.60 ng/ml



#26 AR-1254-1


R.T.: 5.870 min
Delta R.T.: -0.007 min
Response: 238672
Conc: 209.60 ng/ml



#27 AR-1254-2


R.T.: 7.200 min
Delta R.T.: -0.008 min
Response: 209407
Conc: 169.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.030 min
Delta R.T.: -0.007 min
Response: 213706
Conc: 217.95 ng/ml




#28 AR-1254-3

R.T.: 7.582 min
Delta R.T.: -0.007 min
Response: 122540
Conc: 91.51 ng/ml



#28 AR-1254-3


R.T.: 6.467 min
Delta R.T.: 0.011 min
Response: 396635
Conc: 274.21 ng/ml



#29 AR-1254-4


R.T.: 7.877 min
Delta R.T.: -0.007 min
Response: 86906
Conc: 91.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.690 min
Delta R.T.: -0.007 min
Response: 56377
Conc: 67.97 ng/ml




#30 AR-1254-5

R.T.: 8.304 min
Delta R.T.: -0.008 min
Response: 677487
Conc: 637.38 ng/ml



#30 AR-1254-5


R.T.: 7.118 min
Delta R.T.: -0.007 min
Response: 607381
Conc: 521.06 ng/ml



#31 AR-1260-1


R.T.: 7.747 min
Delta R.T.: -0.010 min
Response: 501525
Conc: 508.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.586 min
Delta R.T.: -0.008 min
Response: 517763
Conc: 540.84 ng/ml




#32 AR-1260-2

R.T.: 8.014 min
Delta R.T.: -0.008 min
Response: 575541
Conc: 487.01 ng/ml



#32 AR-1260-2


R.T.: 6.785 min
Delta R.T.: -0.008 min
Response: 585976
Conc: 540.07 ng/ml



#33 AR-1260-3


R.T.: 8.377 min
Delta R.T.: -0.009 min
Response: 446836
Conc: 492.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.938 min
Delta R.T.: -0.009 min
Response: 539127
Conc: 537.31 ng/ml




#34 AR-1260-4

R.T.: 8.608 min
Delta R.T.: -0.008 min
Response: 524515
Conc: 485.44 ng/ml



#34 AR-1260-4


R.T.: 7.420 min
Delta R.T.: -0.008 min
Response: 429149
Conc: 523.23 ng/ml



#35 AR-1260-5


R.T.: 8.923 min
Delta R.T.: -0.008 min
Response: 997331
Conc: 478.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.671 min
Delta R.T.: -0.008 min
Response: 909265
Conc: 520.99 ng/ml




#36 AR-1262-1

R.T.: 8.377 min
Delta R.T.: -0.007 min
Response: 446836
Conc: 373.98 ng/ml



#36 AR-1262-1


R.T.: 7.118 min
Delta R.T.: -0.006 min
Response: 607381
Conc: 1025.37 ng/ml



#37 AR-1262-2


R.T.: 8.923 min
Delta R.T.: -0.007 min
Response: 997331
Conc: 472.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.671 min
Delta R.T.: -0.007 min
Response: 909265
Conc: 549.33 ng/ml




#38 AR-1262-3

R.T.: 9.221 min
Delta R.T.: -0.005 min
Response: 238026
Conc: 235.80 ng/ml



#38 AR-1262-3


R.T.: 7.956 min
Delta R.T.: -0.006 min
Response: 227373
Conc: 315.66 ng/ml



#39 AR-1262-4


R.T.: 9.289 min
Delta R.T.: -0.025 min
Response: 457746
Conc: 685.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.019 min
Delta R.T.: -0.007 min
Response: 688678
Conc: 535.58 ng/ml




#40 AR-1262-5

R.T.: 9.941 min
Delta R.T.: -0.010 min
Response: 322943
Conc: 392.07 ng/ml



#40 AR-1262-5


R.T.: 8.519 min
Delta R.T.: -0.006 min
Response: 249279
Conc: 415.17 ng/ml



#41 AR-1268-1


R.T.: 9.221 min
Delta R.T.: -0.003 min
Response: 238026
Conc: 87.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.956 min
Delta R.T.: -0.006 min
Response: 227373
Conc: 109.16 ng/ml




#42 AR-1268-2

R.T.: 9.289 min
Delta R.T.: -0.027 min
Response: 457746
Conc: 177.90 ng/ml



#42 AR-1268-2


R.T.: 8.019 min
Delta R.T.: -0.008 min
Response: 688678
Conc: 361.43 ng/ml



#43 AR-1268-3


R.T.: 9.520 min
Delta R.T.: -0.010 min
Response: 13466
Conc: 6.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.224 min
Delta R.T.: -0.008 min
Response: 12105
Conc: 7.45 ng/ml




#44 AR-1268-4

R.T.: 9.941 min
Delta R.T.: -0.008 min
Response: 322943
Conc: 336.55 ng/ml



#44 AR-1268-4


R.T.: 8.519 min
Delta R.T.: -0.006 min
Response: 249279
Conc: 354.62 ng/ml



#45 AR-1268-5

R.T.: 10.333 min
Delta R.T.: -0.007 min
Response: 93783
Conc: 13.27 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.790 min
Delta R.T.: -0.008 min
Response: 70717
Conc: 15.22 ng/ml