

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
Data File : PP041413.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Nov 2021 21:32
Operator : AJ\MA
Sample : M4844-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
K-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 01:16:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.750	3.755	505041	683966	23.948	24.611
2)	SA Decachlor...	10.646	9.013	371530	265851	17.008	18.545

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
Data File : PP041413.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Nov 2021 21:32
Operator : AJ\MA
Sample : M4844-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
K-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 01:16:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.750 min
Delta R.T.: -0.007 min
Response: 505041
Conc: 23.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
K-1




#1 Tetrachloro-m-xylene

R.T.: 3.755 min
Delta R.T.: -0.005 min
Response: 683966
Conc: 24.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.646 min
Delta R.T.: -0.014 min
Response: 371530
Conc: 17.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.013 min
Delta R.T.: -0.013 min
Response: 265851
Conc: 18.55 ng/ml