

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
 Data File : PP041416.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Nov 2021 22:50
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 10:29:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 16 04:45:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.750	3.755	1148667	1512724	54.468	54.433
2) SA Decachlor...	10.645	9.013	1026054	719796	46.972	50.212
Target Compounds						
21) L5 AR-1248-1	6.066	5.001	199963	239980	538.193	496.774
22) L5 AR-1248-2	6.358	5.263	291640	356535	525.949	525.519
23) L5 AR-1248-3	6.574	5.307	347723	365295	516.385	523.512
24) L5 AR-1248-4	7.002	5.490	385945	409671	517.298	511.678
25) L5 AR-1248-5	7.041	5.912	379044	410957	501.331	567.650



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112921\
Data File : PP041416.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Nov 2021 22:50
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 10:29:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 16 04:45:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.750 min
Delta R.T.: -0.007 min
Response: 1148667
Conc: 54.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.755 min
Delta R.T.: -0.005 min
Response: 1512724
Conc: 54.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.645 min
Delta R.T.: -0.010 min
Response: 1026054
Conc: 46.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.013 min
Delta R.T.: -0.011 min
Response: 719796
Conc: 50.21 ng/ml



#21 AR-1248-1


R.T.: 6.066 min
Delta R.T.: -0.008 min
Response: 199963
Conc: 538.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.001 min
Delta R.T.: -0.008 min
Response: 239980
Conc: 496.77 ng/ml




#22 AR-1248-2

R.T.: 6.358 min
Delta R.T.: -0.008 min
Response: 291640
Conc: 525.95 ng/ml



#22 AR-1248-2


R.T.: 5.263 min
Delta R.T.: -0.008 min
Response: 356535
Conc: 525.52 ng/ml



#23 AR-1248-3


R.T.: 6.574 min
Delta R.T.: -0.008 min
Response: 347723
Conc: 516.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.307 min
Delta R.T.: -0.007 min
Response: 365295
Conc: 523.51 ng/ml




#24 AR-1248-4

R.T.: 7.002 min
Delta R.T.: -0.009 min
Response: 385945
Conc: 517.30 ng/ml



#24 AR-1248-4


R.T.: 5.490 min
Delta R.T.: -0.008 min
Response: 409671
Conc: 511.68 ng/ml



#25 AR-1248-5

R.T.: 7.041 min
Delta R.T.: -0.009 min
Response: 379044
Conc: 501.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.912 min
Delta R.T.: -0.009 min
Response: 410957
Conc: 567.65 ng/ml