

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112923\
 Data File : PP061973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2023 14:43
 Operator : YP\AJ
 Sample : 05546-03 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 1147

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/30/2023
 Supervised By :Ankita Jodhani 11/30/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 29 15:16:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.724	4977854	6996872	3.781	3.888m
2) SA Decachlor...	10.244	8.850	3150548	5756204	3.053	2.789

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112923\
Data File : PP061973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2023 14:43
Operator : YP\AJ
Sample : 05546-03 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
1147


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/30/2023
Supervised By :Ankita Jodhani 11/30/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 29 15:16:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 4977854
Conc: 3.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
1147


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/30/2023
Supervised By :Ankita Jodhani 11/30/2023




#1 Tetrachloro-m-xylene

R.T.: 3.724 min
Delta R.T.: -0.002 min
Response: 6996872
Conc: 3.89 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: -0.005 min
Response: 3150548
Conc: 3.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.005 min
Response: 5756204
Conc: 2.79 ng/ml