

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112923\
 Data File : PP061991.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2023 20:16
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 00:31:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.723	70068050	91905583	53.225	51.076
2) SA Decachlor...	10.243	8.850	52500051	98704694	50.872	47.823
Target Compounds						
3) L1 AR-1016-1	5.628	4.837	26362582	33648624	532.017	507.936
4) L1 AR-1016-2	5.650	4.856	38638916	47548128	517.959	514.687
5) L1 AR-1016-3	5.712	5.037	25604524	26083188	520.279	514.162
6) L1 AR-1016-4	5.811	5.079	20228463	21502517	532.204	471.941
7) L1 AR-1016-5	6.107	5.296	20749762	31130976	539.565	538.495
31) L7 AR-1260-1	7.239	6.348	37614536	58746007	550.652	504.784
32) L7 AR-1260-2	7.497	6.538	40492219	69334446	537.594	508.866
33) L7 AR-1260-3	7.858	6.694	27531263	65800306	531.428	502.899
34) L7 AR-1260-4	8.084	7.173	31885920	49504325	522.488	501.983
35) L7 AR-1260-5	8.407	7.416	52796468	109.6E6	526.496	511.854



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP112923\
Data File : PP061991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2023 20:16
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 00:31:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 70068050
Conc: 53.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.723 min
Delta R.T.: -0.003 min
Response: 91905583
Conc: 51.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: -0.005 min
Response: 52500051
Conc: 50.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: -0.005 min
Response: 98704694
Conc: 47.82 ng/ml



#3 AR-1016-1


R.T.: 5.628 min
Delta R.T.: -0.001 min
Response: 26362582
Conc: 532.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.837 min
Delta R.T.: -0.003 min
Response: 33648624
Conc: 507.94 ng/ml




#4 AR-1016-2

R.T.: 5.650 min
Delta R.T.: -0.002 min
Response: 38638916
Conc: 517.96 ng/ml



#4 AR-1016-2


R.T.: 4.856 min
Delta R.T.: -0.003 min
Response: 47548128
Conc: 514.69 ng/ml



#5 AR-1016-3


R.T.: 5.712 min
Delta R.T.: -0.002 min
Response: 25604524
Conc: 520.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.037 min
Delta R.T.: -0.003 min
Response: 26083188
Conc: 514.16 ng/ml




#6 AR-1016-4

R.T.: 5.811 min
Delta R.T.: -0.002 min
Response: 20228463
Conc: 532.20 ng/ml



#6 AR-1016-4


R.T.: 5.079 min
Delta R.T.: -0.003 min
Response: 21502517
Conc: 471.94 ng/ml



#7 AR-1016-5


R.T.: 6.107 min
Delta R.T.: -0.002 min
Response: 20749762
Conc: 539.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.296 min
Delta R.T.: -0.003 min
Response: 31130976
Conc: 538.50 ng/ml




#31 AR-1260-1

R.T.: 7.239 min
Delta R.T.: -0.002 min
Response: 37614536
Conc: 550.65 ng/ml



#31 AR-1260-1


R.T.: 6.348 min
Delta R.T.: -0.003 min
Response: 58746007
Conc: 504.78 ng/ml



#32 AR-1260-2


R.T.: 7.497 min
Delta R.T.: -0.002 min
Response: 40492219
Conc: 537.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.538 min
Delta R.T.: -0.003 min
Response: 69334446
Conc: 508.87 ng/ml




#33 AR-1260-3

R.T.: 7.858 min
Delta R.T.: -0.003 min
Response: 27531263
Conc: 531.43 ng/ml



#33 AR-1260-3


R.T.: 6.694 min
Delta R.T.: -0.003 min
Response: 65800306
Conc: 502.90 ng/ml



#34 AR-1260-4


R.T.: 8.084 min
Delta R.T.: -0.003 min
Response: 31885920
Conc: 522.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.173 min
Delta R.T.: -0.003 min
Response: 49504325
Conc: 501.98 ng/ml



#35 AR-1260-5

R.T.: 8.407 min
Delta R.T.: -0.003 min
Response: 52796468
Conc: 526.50 ng/ml



#35 AR-1260-5

R.T.: 7.416 min
Delta R.T.: -0.003 min
Response: 109601679
Conc: 511.85 ng/ml