

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113020\
 Data File : PP031723.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2020 11:19
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 12:28:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.790	4.064	1221507	1095884	25.226	22.870
2)	SA Decachlor...	10.670	9.369	835454	878735	24.598	23.732
Target Compounds							
9)	L2 AR-1221-2	0.000	4.418	0	12082	N.D.	31.320 #
30)	L6 AR-1254-5	0.000	7.432	0	3052	N.D.	1.131 #
32)	L7 AR-1260-2	0.000	7.112	0	5905	N.D.	2.425 #
35)	L7 AR-1260-5	0.000	7.988	0	5031	N.D.	1.109 #
37)	L8 AR-1262-2	0.000	7.988	0	5031	N.D.	1.036 #
38)	L8 AR-1262-3	0.000	8.268	0	4447	N.D.	2.261 #
39)	L8 AR-1262-4	0.000	8.337	0	2147	N.D.	0.592 #
40)	L8 AR-1262-5	0.000	8.838	0	3817	N.D.	2.351 #
41)	L9 AR-1268-1	0.000	8.268	0	4447	N.D.	0.804 #
42)	L9 AR-1268-2	0.000	8.337	0	2147	N.D.	0.397 #
43)	L9 AR-1268-3	0.000	8.542	0	1485	N.D.	0.324 #
44)	L9 AR-1268-4	0.000	8.838	0	3817	N.D.	2.094 #
45)	L9 AR-1268-5	0.000	9.126	0	4120	N.D.	0.298 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113020\
Data File : PP031723.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2020 11:19
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 12:28:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 1221507
Conc: 25.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.064 min
Delta R.T.: 0.003 min
Response: 1095884
Conc: 22.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.670 min
Delta R.T.: 0.005 min
Response: 835454
Conc: 24.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.369 min
Delta R.T.: -0.002 min
Response: 878735
Conc: 23.73 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 5.133 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.418 min
Delta R.T.: 0.000 min
Response: 12082
Conc: 31.32 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 8.328 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.432 min
Delta R.T.: -0.004 min
Response: 3052
Conc: 1.13 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 8.040 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.112 min
Delta R.T.: 0.009 min
Response: 5905
Conc: 2.42 ng/ml




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T.: 8.948 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 5031
Conc: 1.11 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 8.948 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.988 min
Delta R.T.: 0.002 min
Response: 5031
Conc: 1.04 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 9.246 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 8.268 min
Delta R.T.: -0.004 min
Response: 4447
Conc: 2.26 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.334 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.337 min
Delta R.T.: 0.001 min
Response: 2147
Conc: 0.59 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.961 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.838 min
Delta R.T.: 0.005 min
Response: 3817
Conc: 2.35 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 9.244 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.268 min
Delta R.T.: -0.005 min
Response: 4447
Conc: 0.80 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 9.336 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.337 min
Delta R.T.: 0.000 min
Response: 2147
Conc: 0.40 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.550 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.542 min
Delta R.T.: -0.003 min
Response: 1485
Conc: 0.32 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.960 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.838 min
Delta R.T.: 0.005 min
Response: 3817
Conc: 2.09 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.347 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.126 min
Delta R.T.: 0.007 min
Response: 4120
Conc: 0.30 ng/ml