

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113020\
 Data File : PP031733.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2020 14:33
 Operator : DD\AJ
 Sample : L4903-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 15:40:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.792	4.066	1305325	1195301	26.957	24.944
2) SA Decachlor...	10.672	9.373	852709	855823	25.106	23.113
Target Compounds						
6) L1 AR-1016-4	0.000	5.565f	0	5144	N.D.	5.893 #
7) L1 AR-1016-5	0.000	5.801	0	10582	N.D.	9.304 #
8) L2 AR-1221-1	5.032	4.346f	120442	39182	221.463	77.229 #
9) L2 AR-1221-2	0.000	4.421	0	42599	N.D.	110.432 #
15) L3 AR-1232-5	0.000	5.801	0	10582	N.D.	36.518 #
22) L5 AR-1248-2	0.000	5.565f	0	5144	N.D.	4.308 #
24) L5 AR-1248-4	0.000	5.801	0	10582	N.D.	6.907 #
31) L7 AR-1260-1	7.755f	0.000	312956	0	178.827	N.D. #
32) L7 AR-1260-2	0.000	7.077f	0	10769	N.D.	4.422 #
34) L7 AR-1260-4	0.000	7.748	0	5911	N.D.	3.111 #
35) L7 AR-1260-5	0.000	7.981	0	13573	N.D.	2.991 #
37) L8 AR-1262-2	0.000	7.981	0	13573	N.D.	2.794 #
38) L8 AR-1262-3	0.000	8.282	0	11800	N.D.	5.999 #
39) L8 AR-1262-4	0.000	8.350	0	10082	N.D.	2.781 #
40) L8 AR-1262-5	0.000	8.837	0	3936	N.D.	2.425 #
41) L9 AR-1268-1	0.000	8.282	0	11800	N.D.	2.133 #
42) L9 AR-1268-2	0.000	8.350	0	10082	N.D.	1.865 #
43) L9 AR-1268-3	0.000	8.553	0	4072	N.D.	0.888 #
44) L9 AR-1268-4	0.000	8.837	0	3936	N.D.	2.159 #
45) L9 AR-1268-5	0.000	9.130	0	11638	N.D.	0.840 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113020\
Data File : PP031733.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2020 14:33
Operator : DD\AJ
Sample : L4903-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 15:40:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.792 min
Delta R.T.: 0.002 min
Response: 1305325
Conc: 26.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.066 min
Delta R.T.: 0.004 min
Response: 1195301
Conc: 24.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.672 min
Delta R.T.: 0.007 min
Response: 852709
Conc: 25.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.373 min
Delta R.T.: 0.002 min
Response: 855823
Conc: 23.11 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 6.289 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#6 AR-1016-4

R.T.: 5.565 min
Delta R.T.: -0.017 min
Response: 5144
Conc: 5.89 ng/ml




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T.: 6.603 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.801 min
Delta R.T.: -0.010 min
Response: 10582
Conc: 9.30 ng/ml



#8 AR-1221-1


R.T.: 5.032 min
Delta R.T.: -0.004 min
Response: 120442
Conc: 221.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 4.346 min
Delta R.T.: 0.027 min
Response: 39182
Conc: 77.23 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.133 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.421 min
Delta R.T.: 0.003 min
Response: 42599
Conc: 110.43 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 6.389 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.801 min
Delta R.T.: -0.006 min
Response: 10582
Conc: 36.52 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T.: 6.389 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 5.565 min
Delta R.T.: -0.016 min
Response: 5144
Conc: 4.31 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 7.029 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#24 AR-1248-4

R.T.: 5.801 min
Delta R.T.: -0.010 min
Response: 10582
Conc: 6.91 ng/ml




#31 AR-1260-1

R.T.: 7.755 min
Delta R.T.: -0.020 min
Response: 312956
Conc: 178.83 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.907 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 8.040 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.077 min
Delta R.T.: -0.026 min
Response: 10769
Conc: 4.42 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 8.634 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 7.748 min
Delta R.T.: 0.007 min
Response: 5911
Conc: 3.11 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 8.948 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.981 min
Delta R.T.: -0.006 min
Response: 13573
Conc: 2.99 ng/ml




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 8.948 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 7.981 min
Delta R.T.: -0.005 min
Response: 13573
Conc: 2.79 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 9.246 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.282 min
Delta R.T.: 0.010 min
Response: 11800
Conc: 6.00 ng/ml




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 9.334 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.350 min
Delta R.T.: 0.014 min
Response: 10082
Conc: 2.78 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 9.961 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#40 AR-1262-5

R.T.: 8.837 min
Delta R.T.: 0.004 min
Response: 3936
Conc: 2.42 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.244 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.282 min
Delta R.T.: 0.009 min
Response: 11800
Conc: 2.13 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 9.336 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#42 AR-1268-2

R.T.: 8.350 min
Delta R.T.: 0.013 min
Response: 10082
Conc: 1.86 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 9.550 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.553 min
Delta R.T.: 0.009 min
Response: 4072
Conc: 0.89 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 9.960 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#44 AR-1268-4

R.T.: 8.837 min
Delta R.T.: 0.004 min
Response: 3936
Conc: 2.16 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.347 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.130 min
Delta R.T.: 0.010 min
Response: 11638
Conc: 0.84 ng/ml