

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 22:29
 Operator : AJ\MA
 Sample : PP113021ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 23:33:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	22.549	23.849 E3	-5.8	100	0.00
2 SA Decachlorobiphenyl	21.533	22.313 E3	-3.6	101	0.00
3 L1 AR-1016-1	800.137	818.678	-2.3	98	0.00
4 L1 AR-1016-2	1.191	1.246 E3	-4.6	100	0.00
5 L1 AR-1016-3	745.178	796.798	-6.9	100	0.00
6 L1 AR-1016-4	607.067	644.250	-6.1	100	0.00
7 L1 AR-1016-5	585.998	612.813	-4.6	100	0.00
31 L7 AR-1260-1	0.987	1.057 E3	-7.1	102	0.00
32 L7 AR-1260-2	1.256	1.255 E3	0.1	100	0.00
33 L7 AR-1260-3	905.525	956.055	-5.6	100	0.00
34 L7 AR-1260-4	1.070	1.120 E3	-4.7	100	0.00
35 L7 AR-1260-5	2.072	2.149 E3	-3.7	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	30.376	31.358 E3	-3.2	99	0.00
2 SA Decachlorobiphenyl	15.455	15.455 E3	0.0	99	0.00
3 L1 AR-1016-1	929.037	940.162	-1.2	100	0.00
4 L1 AR-1016-2	1.420	1.426 E3	-0.4	98	0.00
5 L1 AR-1016-3	802.876	798.101	0.6	99	0.00
6 L1 AR-1016-4	630.121	640.894	-1.7	100	0.00
7 L1 AR-1016-5	743.858	799.787	-7.5	101	0.00
31 L7 AR-1260-1	1.145	1.163 E3	-1.6	100	0.00
32 L7 AR-1260-2	1.313	1.302 E3	0.8	98	0.00
33 L7 AR-1260-3	1.338	1.256 E3	6.1	98	0.00
34 L7 AR-1260-4	0.985	0.989 E3	-0.4	99	0.00
35 L7 AR-1260-5	2.067	2.079 E3	-0.6	99	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	262.841	0.000	100.0#	0#	-5.00#
9 L2 AR-1221-2	187.406	0.000	100.0#	0#	-5.10#
10 L2 AR-1221-3	623.586	0.000	100.0#	0#	-5.18#
11 L3 AR-1232-1	516.749	0.000	100.0#	0#	-5.18#
12 L3 AR-1232-2	257.323	0.000	100.0#	0#	-5.77#
13 L3 AR-1232-3	494.861	0.000	100.0#	0#	-6.09#
14 L3 AR-1232-4	245.499	0.000	100.0#	0#	-6.26#
15 L3 AR-1232-5	160.986	0.000	100.0#	0#	-6.36#
16 L4 AR-1242-1	599.741	0.000	100.0#	0#	-6.06#
17 L4 AR-1242-2	918.124	0.000	100.0#	0#	-6.09#
18 L4 AR-1242-3	580.565	0.000	100.0#	0#	-6.15#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 22:29
 Operator : AJ\MA
 Sample : PP113021ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 23:33:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	468.606	0.000	100.0#	0#	-6.26#
20 L4 AR-1242-5	480.558	0.000	100.0#	0#	-7.04#
21 L5 AR-1248-1	437.774	0.000	100.0#	0#	-6.06#
22 L5 AR-1248-2	620.157	0.000	100.0#	0#	-6.36#
23 L5 AR-1248-3	747.102	0.000	100.0#	0#	-6.57#
24 L5 AR-1248-4	837.643	0.000	100.0#	0#	-7.00#
25 L5 AR-1248-5	832.153	0.000	100.0#	0#	-7.04#
26 L6 AR-1254-1	848.705	0.000	100.0#	0#	-6.97#
27 L6 AR-1254-2	1.317	0.000 E3	100.0#	0#	-7.20#
28 L6 AR-1254-3	1.424	0.000 E3	100.0#	0#	-7.58#
29 L6 AR-1254-4	1.037	0.000 E3	100.0#	0#	-7.88#
30 L6 AR-1254-5	1.122	0.000 E3	100.0#	0#	-8.30#
36 L8 AR-1262-1	1.270	0.000 E3	100.0#	0#	-8.37#
37 L8 AR-1262-2	2.259	0.000 E3	100.0#	0#	-8.92#
38 L8 AR-1262-3	1.085	0.000 E3	100.0#	0#	-9.21#
39 L8 AR-1262-4	684.434	0.000	100.0#	0#	-9.30#
40 L8 AR-1262-5	894.717	0.000	100.0#	0#	-9.94#
41 L9 AR-1268-1	2.707	0.000 E3	100.0#	0#	-9.21#
42 L9 AR-1268-2	2.619	0.000 E3	100.0#	0#	-9.31#
43 L9 AR-1268-3	2.216	0.000 E3	100.0#	0#	-9.52#
44 L9 AR-1268-4	980.550	0.000	100.0#	0#	-9.94#
45 L9 AR-1268-5	7.228	0.000 E3	100.0#	0#	-10.33#

Signal #2

8 L2 AR-1221-1	350.935	0.000	100.0#	0#	-4.01#
9 L2 AR-1221-2	275.291	0.000	100.0#	0#	-4.11#
10 L2 AR-1221-3	867.852	0.000	100.0#	0#	-4.19#
11 L3 AR-1232-1	740.750	0.000	100.0#	0#	-4.19#
12 L3 AR-1232-2	620.589	0.000	100.0#	0#	-5.02#
13 L3 AR-1232-3	323.896	0.000	100.0#	0#	-5.21#
14 L3 AR-1232-4	285.479	0.000	100.0#	0#	-5.31#
15 L3 AR-1232-5	307.701	0.000	100.0#	0#	-5.49#
16 L4 AR-1242-1	736.447	0.000	100.0#	0#	-5.00#
17 L4 AR-1242-2	1.114	0.000 E3	100.0#	0#	-5.02#
18 L4 AR-1242-3	614.187	0.000	100.0#	0#	-5.21#
19 L4 AR-1242-4	586.843	0.000	100.0#	0#	-5.31#
20 L4 AR-1242-5	650.829	0.000	100.0#	0#	-5.87#
21 L5 AR-1248-1	546.507	0.000	100.0#	0#	-5.00#
22 L5 AR-1248-2	814.141	0.000	100.0#	0#	-5.26#
23 L5 AR-1248-3	846.869	0.000	100.0#	0#	-5.31#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 22:29
 Operator : AJ\MA
 Sample : PP113021ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 23:33:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	975.219	0.000	100.0#	0#	-5.49#
25 L5 AR-1248-5	895.381	0.000	100.0#	0#	-5.91#
26 L6 AR-1254-1	1.392	0.000 E3	100.0#	0#	-5.87#
27 L6 AR-1254-2	1.194	0.000 E3	100.0#	0#	-6.03#
28 L6 AR-1254-3	1.770	0.000 E3	100.0#	0#	-6.45#
29 L6 AR-1254-4	1.031	0.000 E3	100.0#	0#	-6.69#
30 L6 AR-1254-5	1.458	0.000 E3	100.0#	0#	-7.12#
36 L8 AR-1262-1	747.415	0.000	100.0#	0#	-7.11#
37 L8 AR-1262-2	2.110	0.000 E3	100.0#	0#	-7.67#
38 L8 AR-1262-3	891.168	0.000	100.0#	0#	-7.95#
39 L8 AR-1262-4	1.606	0.000 E3	100.0#	0#	-8.02#
40 L8 AR-1262-5	725.979	0.000	100.0#	0#	-8.51#
41 L9 AR-1268-1	2.398	0.000 E3	100.0#	0#	-7.95#
42 L9 AR-1268-2	2.265	0.000 E3	100.0#	0#	-8.02#
43 L9 AR-1268-3	1.878	0.000 E3	100.0#	0#	-8.22#
44 L9 AR-1268-4	801.418	0.000	100.0#	0#	-8.51#
45 L9 AR-1268-5	5.115	0.000 E3	100.0#	0#	-8.79#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0