

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041449.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 00:42
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 02:14:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	22.549	25.017 E3	-10.9	105	0.00
2 SA	Decachlorobiphenyl	21.533	23.908 E3	-11.0	108	0.00
21 L5	AR-1248-1	437.774	448.548	-2.5	98	0.00
22 L5	AR-1248-2	620.157	641.064	-3.4	102	0.00
23 L5	AR-1248-3	747.102	782.706	-4.8	102	0.00
24 L5	AR-1248-4	837.643	898.921	-7.3	105	0.00
25 L5	AR-1248-5	832.153	942.935	-13.3	110	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	30.376	34.862 E3	-14.8	110	0.00
2 SA	Decachlorobiphenyl	15.455	16.807 E3	-8.7	107	0.00
21 L5	AR-1248-1	546.507	546.965	-0.1	97	0.00
22 L5	AR-1248-2	814.141	830.244	-2.0	102	0.00
23 L5	AR-1248-3	846.869	842.645	0.5	100	0.00
24 L5	AR-1248-4	975.219	936.581	4.0	96	0.00
25 L5	AR-1248-5	895.381	908.318	-1.4	103	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1	AR-1016-1	800.137	0.000	100.0#	0#	-6.06#
4 L1	AR-1016-2	1.191	0.000 E3	100.0#	0#	-6.09#
5 L1	AR-1016-3	745.178	0.000	100.0#	0#	-6.15#
6 L1	AR-1016-4	607.067	0.000	100.0#	0#	-6.26#
7 L1	AR-1016-5	585.998	0.000	100.0#	0#	-6.57#
8 L2	AR-1221-1	262.841	0.000	100.0#	0#	-5.00#
9 L2	AR-1221-2	187.406	0.000	100.0#	0#	-5.10#
10 L2	AR-1221-3	623.586	0.000	100.0#	0#	-5.18#
11 L3	AR-1232-1	516.749	0.000	100.0#	0#	-5.18#
12 L3	AR-1232-2	257.323	0.000	100.0#	0#	-5.77#
13 L3	AR-1232-3	494.861	0.000	100.0#	0#	-6.09#
14 L3	AR-1232-4	245.499	0.000	100.0#	0#	-6.26#
15 L3	AR-1232-5	160.986	0.000	100.0#	0#	-6.36#
16 L4	AR-1242-1	599.741	0.000	100.0#	0#	-6.06#
17 L4	AR-1242-2	918.124	0.000	100.0#	0#	-6.09#
18 L4	AR-1242-3	580.565	0.000	100.0#	0#	-6.15#
19 L4	AR-1242-4	468.606	0.000	100.0#	0#	-6.26#
20 L4	AR-1242-5	480.558	0.000	100.0#	0#	-7.04#
26 L6	AR-1254-1	848.705	0.000	100.0#	0#	-6.97#
27 L6	AR-1254-2	1.317	0.000 E3	100.0#	0#	-7.20#
28 L6	AR-1254-3	1.424	0.000 E3	100.0#	0#	-7.58#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041449.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 00:42
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 02:14:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	1.037	0.000 E3	100.0#	0#	-7.88#
30 L6 AR-1254-5	1.122	0.000 E3	100.0#	0#	-8.30#
31 L7 AR-1260-1	0.987	0.000 E3	100.0#	0#	-7.75#
32 L7 AR-1260-2	1.256	0.000 E3	100.0#	0#	-8.01#
33 L7 AR-1260-3	905.525	0.000	100.0#	0#	-8.38#
34 L7 AR-1260-4	1.070	0.000 E3	100.0#	0#	-8.61#
35 L7 AR-1260-5	2.072	0.000 E3	100.0#	0#	-8.92#
36 L8 AR-1262-1	1.270	0.000 E3	100.0#	0#	-8.37#
37 L8 AR-1262-2	2.259	0.000 E3	100.0#	0#	-8.92#
38 L8 AR-1262-3	1.085	0.000 E3	100.0#	0#	-9.21#
39 L8 AR-1262-4	684.434	0.000	100.0#	0#	-9.30#
40 L8 AR-1262-5	894.717	0.000	100.0#	0#	-9.94#
41 L9 AR-1268-1	2.707	0.000 E3	100.0#	0#	-9.21#
42 L9 AR-1268-2	2.619	0.000 E3	100.0#	0#	-9.31#
43 L9 AR-1268-3	2.216	0.000 E3	100.0#	0#	-9.52#
44 L9 AR-1268-4	980.550	0.000	100.0#	0#	-9.94#
45 L9 AR-1268-5	7.228	0.000 E3	100.0#	0#	-10.33#

Signal #2

3 L1 AR-1016-1	929.037	0.000	100.0#	0#	-5.00#
4 L1 AR-1016-2	1.420	0.000 E3	100.0#	0#	-5.02#
5 L1 AR-1016-3	802.876	0.000	100.0#	0#	-5.21#
6 L1 AR-1016-4	630.121	0.000	100.0#	0#	-5.26#
7 L1 AR-1016-5	743.858	0.000	100.0#	0#	-5.49#
8 L2 AR-1221-1	350.935	0.000	100.0#	0#	-4.01#
9 L2 AR-1221-2	275.291	0.000	100.0#	0#	-4.11#
10 L2 AR-1221-3	867.852	0.000	100.0#	0#	-4.19#
11 L3 AR-1232-1	740.750	0.000	100.0#	0#	-4.19#
12 L3 AR-1232-2	620.589	0.000	100.0#	0#	-5.02#
13 L3 AR-1232-3	323.896	0.000	100.0#	0#	-5.21#
14 L3 AR-1232-4	285.479	0.000	100.0#	0#	-5.31#
15 L3 AR-1232-5	307.701	0.000	100.0#	0#	-5.49#
16 L4 AR-1242-1	736.447	0.000	100.0#	0#	-5.00#
17 L4 AR-1242-2	1.114	0.000 E3	100.0#	0#	-5.02#
18 L4 AR-1242-3	614.187	0.000	100.0#	0#	-5.21#
19 L4 AR-1242-4	586.843	0.000	100.0#	0#	-5.31#
20 L4 AR-1242-5	650.829	0.000	100.0#	0#	-5.87#
26 L6 AR-1254-1	1.392	0.000 E3	100.0#	0#	-5.87#
27 L6 AR-1254-2	1.194	0.000 E3	100.0#	0#	-6.03#
28 L6 AR-1254-3	1.770	0.000 E3	100.0#	0#	-6.45#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041449.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 00:42
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 02:14:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	1.031	0.000 E3	100.0#	0#	-6.69#
30 L6 AR-1254-5	1.458	0.000 E3	100.0#	0#	-7.12#
31 L7 AR-1260-1	1.145	0.000 E3	100.0#	0#	-6.58#
32 L7 AR-1260-2	1.313	0.000 E3	100.0#	0#	-6.78#
33 L7 AR-1260-3	1.338	0.000 E3	100.0#	0#	-6.93#
34 L7 AR-1260-4	0.985	0.000 E3	100.0#	0#	-7.42#
35 L7 AR-1260-5	2.067	0.000 E3	100.0#	0#	-7.67#
36 L8 AR-1262-1	747.415	0.000	100.0#	0#	-7.11#
37 L8 AR-1262-2	2.110	0.000 E3	100.0#	0#	-7.67#
38 L8 AR-1262-3	891.168	0.000	100.0#	0#	-7.95#
39 L8 AR-1262-4	1.606	0.000 E3	100.0#	0#	-8.02#
40 L8 AR-1262-5	725.979	0.000	100.0#	0#	-8.51#
41 L9 AR-1268-1	2.398	0.000 E3	100.0#	0#	-7.95#
42 L9 AR-1268-2	2.265	0.000 E3	100.0#	0#	-8.02#
43 L9 AR-1268-3	1.878	0.000 E3	100.0#	0#	-8.22#
44 L9 AR-1268-4	801.418	0.000	100.0#	0#	-8.51#
45 L9 AR-1268-5	5.115	0.000 E3	100.0#	0#	-8.79#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0