

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041421.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 14:57
 Operator : AJ\MA
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 16:10:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 16:07:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.747	3.751	2306614	2969517	97.028	93.506
2) SA Decachlor...	10.644	9.012	2047206	1384485	92.656	88.344
Target Compounds						
3) L1 AR-1016-1	6.063	4.999	770741	837122	925.622	886.293
4) L1 AR-1016-2	6.087	5.017	1163044	1311657	930.577	903.627
5) L1 AR-1016-3	6.152	5.208	733247	720352	921.760	893.482
6) L1 AR-1016-4	6.258	5.262	604070	567996	937.050	886.475
7) L1 AR-1016-5	6.572	5.489	576888	707860	945.688	890.780
31) L7 AR-1260-1	7.745	6.583	965417	1028558	934.706	885.332
32) L7 AR-1260-2	8.012	6.782	1145912	1175952	917.260	883.383
33) L7 AR-1260-3	8.375	6.935	881878	1126246	925.382	876.921
34) L7 AR-1260-4	8.606	7.418	1043484	882820	931.511	882.620
35) L7 AR-1260-5	8.921	7.668	2022209	1901289	946.951	905.549



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
Data File : PP041421.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2021 14:57
Operator : AJ\MA
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 16:10:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 16:07:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 2306614
Conc: 97.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.751 min
Delta R.T.: 0.000 min
Response: 2969517
Conc: 93.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.644 min
Delta R.T.: -0.001 min
Response: 2047206
Conc: 92.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.012 min
Delta R.T.: 0.000 min
Response: 1384485
Conc: 88.34 ng/ml



#3 AR-1016-1


R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 770741
Conc: 925.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 837122
Conc: 886.29 ng/ml




#4 AR-1016-2

R.T.: 6.087 min
Delta R.T.: 0.000 min
Response: 1163044
Conc: 930.58 ng/ml



#4 AR-1016-2


R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 1311657
Conc: 903.63 ng/ml



#5 AR-1016-3


R.T.: 6.152 min
Delta R.T.: 0.000 min
Response: 733247
Conc: 921.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.208 min
Delta R.T.: 0.000 min
Response: 720352
Conc: 893.48 ng/ml




#6 AR-1016-4

R.T.: 6.258 min
Delta R.T.: 0.000 min
Response: 604070
Conc: 937.05 ng/ml



#6 AR-1016-4


R.T.: 5.262 min
Delta R.T.: 0.000 min
Response: 567996
Conc: 886.48 ng/ml



#7 AR-1016-5


R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 576888
Conc: 945.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.489 min
Delta R.T.: 0.000 min
Response: 707860
Conc: 890.78 ng/ml




#31 AR-1260-1

R.T.: 7.745 min
Delta R.T.: 0.000 min
Response: 965417
Conc: 934.71 ng/ml



#31 AR-1260-1


R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 1028558
Conc: 885.33 ng/ml



#32 AR-1260-2


R.T.: 8.012 min
Delta R.T.: 0.000 min
Response: 1145912
Conc: 917.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.782 min
Delta R.T.: 0.000 min
Response: 1175952
Conc: 883.38 ng/ml




#33 AR-1260-3

R.T.: 8.375 min
Delta R.T.: 0.000 min
Response: 881878
Conc: 925.38 ng/ml



#33 AR-1260-3


R.T.: 6.935 min
Delta R.T.: 0.000 min
Response: 1126246
Conc: 876.92 ng/ml



#34 AR-1260-4


R.T.: 8.606 min
Delta R.T.: 0.000 min
Response: 1043484
Conc: 931.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.418 min
Delta R.T.: 0.001 min
Response: 882820
Conc: 882.62 ng/ml



#35 AR-1260-5

R.T.: 8.921 min
Delta R.T.: 0.000 min
Response: 2022209
Conc: 946.95 ng/ml



#35 AR-1260-5

R.T.: 7.668 min
Delta R.T.: 0.000 min
Response: 1901289
Conc: 905.55 ng/ml