

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041434.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 19:25
 Operator : AJ\MA
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 22:38:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 22:37:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.748	3.753	1828704	2518222	73.622	71.802
2)	SA Decachlor...	10.645	9.013	1679542	1147928	71.081	70.010
Target Compounds							
21)	L5 AR-1248-1	6.064	5.000	326170	379030	711.039	669.131
22)	L5 AR-1248-2	6.357	5.264	452086	564878	721.109	691.116
23)	L5 AR-1248-3	6.573	5.307	554224	593414	721.260	700.941
24)	L5 AR-1248-4	7.002	5.491	614221	685331	717.176	701.203
25)	L5 AR-1248-5	7.041	5.912	619768	622573	721.051	704.767



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
Data File : PP041434.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2021 19:25
Operator : AJ\MA
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 22:38:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 22:37:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 1828704
Conc: 73.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.753 min
Delta R.T.: 0.000 min
Response: 2518222
Conc: 71.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.645 min
Delta R.T.: -0.002 min
Response: 1679542
Conc: 71.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.013 min
Delta R.T.: 0.000 min
Response: 1147928
Conc: 70.01 ng/ml



#21 AR-1248-1


R.T.: 6.064 min
Delta R.T.: 0.000 min
Response: 326170
Conc: 711.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 379030
Conc: 669.13 ng/ml




#22 AR-1248-2

R.T.: 6.357 min
Delta R.T.: 0.000 min
Response: 452086
Conc: 721.11 ng/ml



#22 AR-1248-2


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 564878
Conc: 691.12 ng/ml



#23 AR-1248-3


R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 554224
Conc: 721.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.307 min
Delta R.T.: 0.000 min
Response: 593414
Conc: 700.94 ng/ml




#24 AR-1248-4

R.T.: 7.002 min
Delta R.T.: 0.000 min
Response: 614221
Conc: 717.18 ng/ml



#24 AR-1248-4


R.T.: 5.491 min
Delta R.T.: 0.000 min
Response: 685331
Conc: 701.20 ng/ml



#25 AR-1248-5

R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 619768
Conc: 721.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 622573
Conc: 704.77 ng/ml