

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041440.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 21:06
 Operator : AJ\MA
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 22:54:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 22:53:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.747	3.751	1216852	1697227	50.000	50.000
2) SA Decachlor...	10.642	9.011	1155024	803647	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.969	5.868	434214	677743	500.000	500.000
27) L6 AR-1254-2	7.198	6.028	672126	583132	500.000	500.000
28) L6 AR-1254-3	7.579	6.446	722727	869650	500.000	500.000
29) L6 AR-1254-4	7.874	6.686	534796	510663	500.000	500.000
30) L6 AR-1254-5	8.302	7.115	584164	722524	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
Data File : PP041440.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2021 21:06
Operator : AJ\MA
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 22:54:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 22:53:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 1216852
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.751 min
Delta R.T.: 0.000 min
Response: 1697227
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.642 min
Delta R.T.: 0.000 min
Response: 1155024
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.011 min
Delta R.T.: 0.000 min
Response: 803647
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.969 min
Delta R.T.: 0.000 min
Response: 434214
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500




#26 AR-1254-1

R.T.: 5.868 min
Delta R.T.: 0.000 min
Response: 677743
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 672126
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 6.028 min
Delta R.T.: 0.000 min
Response: 583132
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.579 min
Delta R.T.: 0.000 min
Response: 722727
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.446 min
Delta R.T.: 0.000 min
Response: 869650
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.874 min
Delta R.T.: 0.000 min
Response: 534796
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.686 min
Delta R.T.: 0.000 min
Response: 510663
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 8.302 min
Delta R.T.: 0.000 min
Response: 584164
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500



#30 AR-1254-5

R.T.: 7.115 min
Delta R.T.: 0.000 min
Response: 722524
Conc: 500.00 ng/ml