

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 22:29
 Operator : AJ\MA
 Sample : PP113021ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP113021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 23:33:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.747	3.751	1192460	1567893	52.882	51.616
2) SA Decachlor...	10.643	9.010	1115656	772755	51.813	49.999
Target Compounds						
3) L1 AR-1016-1	6.063	4.999	409339	470081	511.586	505.988
4) L1 AR-1016-2	6.086	5.018	623242	713225	523.087	502.418
5) L1 AR-1016-3	6.151	5.209	398399	399050	534.636	497.026
6) L1 AR-1016-4	6.258	5.262	322125	320447	530.625	508.548
7) L1 AR-1016-5	6.571	5.489	306407	399893	522.880	537.594
31) L7 AR-1260-1	7.744	6.582	528651	581465	535.510	507.635
32) L7 AR-1260-2	8.011	6.781	627280	651082	499.263	496.003
33) L7 AR-1260-3	8.374	6.934	478027	628231	527.901	469.679
34) L7 AR-1260-4	8.604	7.416	559914	494540	523.223	502.251
35) L7 AR-1260-5	8.919	7.667	1074338	1039650	518.509	502.978



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
Data File : PP041445.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2021 22:29
Operator : AJ\MA
Sample : PP113021ICV500
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP113021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 23:33:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 1192460
Conc: 52.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#1 Tetrachloro-m-xylene

R.T.: 3.751 min
Delta R.T.: 0.000 min
Response: 1567893
Conc: 51.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.643 min
Delta R.T.: 0.000 min
Response: 1115656
Conc: 51.81 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.010 min
Delta R.T.: -0.002 min
Response: 772755
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 409339
Conc: 511.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#3 AR-1016-1

R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 470081
Conc: 505.99 ng/ml




#4 AR-1016-2

R.T.: 6.086 min
Delta R.T.: 0.000 min
Response: 623242
Conc: 523.09 ng/ml



#4 AR-1016-2


R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 713225
Conc: 502.42 ng/ml



#5 AR-1016-3


R.T.: 6.151 min
Delta R.T.: 0.000 min
Response: 398399
Conc: 534.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#5 AR-1016-3

R.T.: 5.209 min
Delta R.T.: 0.000 min
Response: 399050
Conc: 497.03 ng/ml




#6 AR-1016-4

R.T.: 6.258 min
Delta R.T.: 0.000 min
Response: 322125
Conc: 530.63 ng/ml



#6 AR-1016-4


R.T.: 5.262 min
Delta R.T.: 0.000 min
Response: 320447
Conc: 508.55 ng/ml



#7 AR-1016-5


R.T.: 6.571 min
Delta R.T.: 0.000 min
Response: 306407
Conc: 522.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#7 AR-1016-5

R.T.: 5.489 min
Delta R.T.: 0.000 min
Response: 399893
Conc: 537.59 ng/ml




#31 AR-1260-1

R.T.: 7.744 min
Delta R.T.: -0.001 min
Response: 528651
Conc: 535.51 ng/ml



#31 AR-1260-1


R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 581465
Conc: 507.63 ng/ml



#32 AR-1260-2


R.T.: 8.011 min
Delta R.T.: -0.001 min
Response: 627280
Conc: 499.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#32 AR-1260-2

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 651082
Conc: 496.00 ng/ml




#33 AR-1260-3

R.T.: 8.374 min
Delta R.T.: -0.001 min
Response: 478027
Conc: 527.90 ng/ml



#33 AR-1260-3


R.T.: 6.934 min
Delta R.T.: 0.000 min
Response: 628231
Conc: 469.68 ng/ml



#34 AR-1260-4


R.T.: 8.604 min
Delta R.T.: -0.002 min
Response: 559914
Conc: 523.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#34 AR-1260-4

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 494540
Conc: 502.25 ng/ml



#35 AR-1260-5

R.T.: 8.919 min
Delta R.T.: -0.002 min
Response: 1074338
Conc: 518.51 ng/ml



#35 AR-1260-5

R.T.: 7.667 min
Delta R.T.: -0.001 min
Response: 1039650
Conc: 502.98 ng/ml