

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
 Data File : PP041448.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Nov 2021 23:19
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP113021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 23:34:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.746	3.751	1224842	1728757	54.318	56.912
2)	SA Decachlor...	10.642	9.010	1154913	800313	53.636	51.782
Target Compounds							
26)	L6 AR-1254-1	6.968	5.867	431501	677541	508.423	486.723
27)	L6 AR-1254-2	7.198	6.027	669973	582839	508.558	488.040
28)	L6 AR-1254-3	7.578	6.446	722482	868390	507.222	490.541
29)	L6 AR-1254-4	7.875	6.685	532857	512553	513.597	497.289
30)	L6 AR-1254-5	8.302	7.114	582513	721644	519.021	494.935



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113021\
Data File : PP041448.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Nov 2021 23:19
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP113021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 23:34:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.746 min
Delta R.T.: -0.001 min
Response: 1224842
Conc: 54.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#1 Tetrachloro-m-xylene

R.T.: 3.751 min
Delta R.T.: 0.000 min
Response: 1728757
Conc: 56.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.642 min
Delta R.T.: -0.002 min
Response: 1154913
Conc: 53.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.010 min
Delta R.T.: -0.002 min
Response: 800313
Conc: 51.78 ng/ml



#26 AR-1254-1


R.T.: 6.968 min
Delta R.T.: -0.002 min
Response: 431501
Conc: 508.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#26 AR-1254-1

R.T.: 5.867 min
Delta R.T.: -0.002 min
Response: 677541
Conc: 486.72 ng/ml




#27 AR-1254-2

R.T.: 7.198 min
Delta R.T.: -0.002 min
Response: 669973
Conc: 508.56 ng/ml



#27 AR-1254-2


R.T.: 6.027 min
Delta R.T.: -0.002 min
Response: 582839
Conc: 488.04 ng/ml



#28 AR-1254-3


R.T.: 7.578 min
Delta R.T.: -0.002 min
Response: 722482
Conc: 507.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021




#28 AR-1254-3

R.T.: 6.446 min
Delta R.T.: -0.002 min
Response: 868390
Conc: 490.54 ng/ml




#29 AR-1254-4

R.T.: 7.875 min
Delta R.T.: -0.002 min
Response: 532857
Conc: 513.60 ng/ml



#29 AR-1254-4


R.T.: 6.685 min
Delta R.T.: -0.003 min
Response: 512553
Conc: 497.29 ng/ml



#30 AR-1254-5

R.T.: 8.302 min
Delta R.T.: -0.002 min
Response: 582513
Conc: 519.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP113021



#30 AR-1254-5

R.T.: 7.114 min
Delta R.T.: -0.003 min
Response: 721644
Conc: 494.93 ng/ml