

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113022\
 Data File : PP053574.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2022 17:11
 Operator : YP\AJ
 Sample : N5809-05
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HOWE-DRUM-C-1129

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/01/2022
 Supervised By :Ankita Jodhani 12/01/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 20:31:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.401	3.646	32751127	29135800	17.037	19.369
2) SA Decachlor...	10.209	8.728	25611069	26621224	18.888	14.203m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113022\
Data File : PP053574.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2022 17:11
Operator : YP\AJ
Sample : N5809-05
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
HOWE-DRUM-C-1129


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/01/2022
Supervised By :Ankita Jodhani 12/01/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 20:31:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.401 min
Delta R.T.: -0.002 min
Response: 32751127
Conc: 17.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
HOWE-DRUM-C-1129


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/01/2022
Supervised By :Ankita Jodhani 12/01/2022




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: -0.003 min
Response: 29135800
Conc: 19.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: 0.007 min
Response: 25611069
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.728 min
Delta R.T.: 0.000 min
Response: 26621224
Conc: 14.20 ng/ml m