

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
 Data File : PP061993.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2023 09:10
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 00:03:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

6)	L1	AR-1016-4	0.000	5.100f	0	335916	N.D.	7.373 #
7)	L1	AR-1016-5	0.000	5.307	0	870126	N.D.	15.051 #
14)	L3	AR-1232-4	0.000	5.121	0	53858	N.D.	2.358 #
15)	L3	AR-1232-5	0.000	5.307	0	870126	N.D.	34.192 #
19)	L4	AR-1242-4	0.000	5.121	0	53858	N.D.	1.193 #
20)	L4	AR-1242-5	0.000	5.680f	0	431208	N.D.	8.240 #
22)	L5	AR-1248-2	0.000	5.100f	0	335916	N.D.	5.318 #
23)	L5	AR-1248-3	0.000	5.121	0	53858	N.D.	0.821 #
24)	L5	AR-1248-4	6.500	5.307	252006	870126	4.972	11.304 #
25)	L5	AR-1248-5	0.000	5.680f	0	431208	N.D.	4.315 #
26)	L6	AR-1254-1	6.500	5.680f	252006	431208	4.492	3.960
28)	L6	AR-1254-3	7.075	0.000	792298	0	9.577	N.D. #
32)	L7	AR-1260-2	7.505	6.519f	90662	804535	1.204	5.905 #
35)	L7	AR-1260-5	0.000	7.414	0	42676	N.D.	0.199 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
Data File : PP061993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2023 09:10
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 00:03:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 4.453 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.726 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 10.248 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T.: 8.855 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 5.813 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#6 AR-1016-4

R.T.: 5.100 min
Delta R.T.: 0.019 min
Response: 335916
Conc: 7.37 ng/ml




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T.: 6.109 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.307 min
Delta R.T.: 0.008 min
Response: 870126
Conc: 15.05 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.813 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#14 AR-1232-4

R.T.: 5.121 min
Delta R.T.: -0.003 min
Response: 53858
Conc: 2.36 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T.: 5.904 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.307 min
Delta R.T.: 0.008 min
Response: 870126
Conc: 34.19 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 5.812 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.121 min
Delta R.T.: -0.003 min
Response: 53858
Conc: 1.19 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.553 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.680 min
Delta R.T.: 0.023 min
Response: 431208
Conc: 8.24 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.903 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#22 AR-1248-2

R.T.: 5.100 min
Delta R.T.: 0.019 min
Response: 335916
Conc: 5.32 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T.: 6.108 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.121 min
Delta R.T.: -0.003 min
Response: 53858
Conc: 0.82 ng/ml



#24 AR-1248-4


R.T.: 6.500 min
Delta R.T.: -0.015 min
Response: 252006
Conc: 4.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#24 AR-1248-4

R.T.: 5.307 min
Delta R.T.: 0.008 min
Response: 870126
Conc: 11.30 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.553 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.680 min
Delta R.T.: 0.023 min
Response: 431208
Conc: 4.31 ng/ml



#26 AR-1254-1


R.T.: 6.500 min
Delta R.T.: 0.012 min
Response: 252006
Conc: 4.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.680 min
Delta R.T.: 0.023 min
Response: 431208
Conc: 3.96 ng/ml




#28 AR-1254-3

R.T.: 7.075 min
Delta R.T.: 0.000 min
Response: 792298
Conc: 9.58 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.217 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 7.505 min
Delta R.T.: 0.006 min
Response: 90662
Conc: 1.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 6.519 min
Delta R.T.: -0.022 min
Response: 804535
Conc: 5.90 ng/ml



#35 AR-1260-5

R.T.: 0.000 min
Exp R.T.: 8.410 min
Response: 0
Conc: N.D.



#35 AR-1260-5

R.T.: 7.414 min
Delta R.T.: -0.005 min
Response: 42676
Conc: 0.20 ng/ml